


LB123763

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.006	-0.006	32	0.000	32					-
0.2	0.200	0.160	-0.040	943	0.0 %	943					-20
2.5	2.500	2.453	-0.047	13493	0.0 %	13493					-2
5.0	5.000	5.065	0.065	27792	0.0 %	27792					-1
7.5	7.500	7.731	0.231	42384	0.0 %	42384					-3
10.0	10.000	9.797	-0.203	53693	0.0 %	53693					-2

LB123763 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Jan 2023 09:31:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	32	-	0.0000	7471B	PPB	23 Jan 2023 10:06:41	S	Std
0.2 - 1	50.2	943	-	0.2000	7471B	PPB	23 Jan 2023 10:11:30	S	Std
2.5 - 1	52.5	13493	-	2.5000	7471B	PPB	23 Jan 2023 10:13:48	S	Std
5.0 - 1	55	27792	-	5.0000	7471B	PPB	23 Jan 2023 10:16:04	S	Std
7.5 - 1	57.5	42384	-	7.5000	7471B	PPB	23 Jan 2023 10:20:53	S	Std
10.0 - 1	50	53693	-	10.0000	7471B	PPB	23 Jan 2023 10:23:09	S	Std
ICV12 - 1	ICV12	21485	3.9130	-	7471B	PPB	23 Jan 2023 10:28:33	U	SMPL
ICB12 - 1	ICB12	-12	-0.0140	-	7471B	PPB	23 Jan 2023 10:30:48	U	SMPL
CCV93 - 1	CCV93	26713	4.8680	-	7471B	PPB	23 Jan 2023 10:33:03	U	SMPL
CCB93 - 1	CCB93	-47	-0.0204	-	7471B	PPB	23 Jan 2023 10:35:18	U	SMPL
CRA - 1	CRA	957	0.1630	-	7471B	PPB	23 Jan 2023 10:37:33	U	SMPL
HighStd - 1	HighStd	53734	9.8041	-	7471B	PPB	23 Jan 2023 10:39:48	U	SMPL
ChkStd - 1	ChkStd	39341	7.1749	-	7471B	PPB	23 Jan 2023 10:42:03	U	SMPL
PB150381BL - 1	PBS	-100	-0.0301	-	7471B	PPB	23 Jan 2023 10:44:19	U	SMPL
PB150381BS - 1	LCSS	19305	3.5147	-	7471B	PPB	23 Jan 2023 10:46:35	U	SMPL
O1229-01 - 1	OK-02-011923	1049	0.1798	-	7471B	PPB	23 Jan 2023 10:48:51	U	SMPL
O1230-01 - 1	BBC	3353	0.6007	-	7471B	PPB	23 Jan 2023 10:51:07	U	SMPL
O1231-01 - 1	BBC	3589	0.6438	-	7471B	PPB	23 Jan 2023 10:53:24	U	SMPL
O1232-01 - 1	B-P-17A	224	0.0291	-	7471B	PPB	23 Jan 2023 10:55:41	U	SMPL
O1232-01DUP - 1	B-P-17ADUP	381	0.0578	-	7471B	PPB	23 Jan 2023 10:57:56	U	SMPL
CCV94 - 1	CCV94	25802	4.7016	-	7471B	PPB	23 Jan 2023 11:00:11	U	SMPL
CCB94 - 1	CCB94	-57	-0.0222	-	7471B	PPB	23 Jan 2023 11:02:27	U	SMPL
O1232-01MS - 1	B-P-17AMS	21674	3.9475	-	7471B	PPB	23 Jan 2023 11:04:42	U	SMPL
O1232-01MSD - 1	B-P-17AMSD	21884	3.9859	-	7471B	PPB	23 Jan 2023 11:06:58	U	SMPL
O1232-02 - 1	B-P-17B	985	0.1681	-	7471B	PPB	23 Jan 2023 11:09:13	U	SMPL
O1232-04 - 1	B-P-18A	1683	0.2956	-	7471B	PPB	23 Jan 2023 11:11:29	U	SMPL
O1232-05 - 1	B-P-18B	381	0.0578	-	7471B	PPB	23 Jan 2023 11:13:45	U	SMPL
O1232-06 - 1	B-P-19A	3591	0.6442	-	7471B	PPB	23 Jan 2023 11:16:02	U	SMPL
O1232-07 - 1	B-P-19B	2992	0.5347	-	7471B	PPB	23 Jan 2023 11:18:18	U	SMPL
O1232-08 - 1	B-P-35A	409	0.0629	-	7471B	PPB	23 Jan 2023 11:20:35	U	SMPL
O1232-09 - 1	B-P-35B	278	0.0390	-	7471B	PPB	23 Jan 2023 11:22:52	U	SMPL
O1232-10 - 1	DUP01	865	0.1462	-	7471B	PPB	23 Jan 2023 11:25:08	U	SMPL
CCV95 - 1	CCV95	25568	4.6589	-	7471B	PPB	23 Jan 2023 11:27:24	U	SMPL
CCB95 - 1	CCB95	-45	-0.0201	-	7471B	PPB	23 Jan 2023 11:29:41	U	SMPL
PB150382BL - 1	PBS	-11	-0.0138	-	7471B	PPB	23 Jan 2023 11:31:57	U	SMPL
PB150382BS - 1	LCSS	18545	3.3759	-	7471B	PPB	23 Jan 2023 11:34:13	U	SMPL
O1233-02 - 1	B-P34A	1415	0.2467	-	7471B	PPB	23 Jan 2023 11:36:28	U	SMPL
O1233-03 - 1	B-P34B	626	0.1025	-	7471B	PPB	23 Jan 2023 11:38:44	U	SMPL
O1233-04 - 1	B-P36A	590	0.0959	-	7471B	PPB	23 Jan 2023 11:40:59	U	SMPL
O1233-05 - 1	B-P36B	259	0.0355	-	7471B	PPB	23 Jan 2023 11:43:16	U	SMPL
O1233-06 - 1	B-P37A	2447	0.4352	-	7471B	PPB	23 Jan 2023 11:45:32	U	SMPL
O1233-07 - 1	B-P37B	1646	0.2889	-	7471B	PPB	23 Jan 2023 11:47:49	U	SMPL
O1233-08 - 1	B-P38A	3730	0.6696	-	7471B	PPB	23 Jan 2023 11:50:06	U	SMPL
O1233-09 - 1	B-P38B	223	0.0289	-	7471B	PPB	23 Jan 2023 11:52:22	U	SMPL
CCV96 - 1	CCV96	26895	4.9013	-	7471B	PPB	23 Jan 2023 11:54:38	U	SMPL
CCB96 - 1	CCB96	-54	-0.0217	-	7471B	PPB	23 Jan 2023 11:56:53	U	SMPL
O1233-11 - 1	B-P38C	225	0.0293	-	7471B	PPB	23 Jan 2023 11:59:09	U	SMPL
O1234-01 - 1	EO-02-11923	1913	0.3376	-	7471B	PPB	23 Jan 2023 12:01:24	U	SMPL
O1237-01 - 1	COMP-1-2-3A	390	0.0594	-	7471B	PPB	23 Jan 2023 12:03:41	U	SMPL
O1239-01 - 1	BORING-3	4650	0.8376	-	7471B	PPB	23 Jan 2023 12:05:56	U	SMPL
O1240-01 - 1	TR-05-011923	568	0.0919	-	7471B	PPB	23 Jan 2023 12:08:12	U	SMPL
O1244-01 - 1	VNJ-223	2202	0.3904	-	7471B	PPB	23 Jan 2023 12:10:29	U	SMPL
O1252-01 - 1	MH-9	709	0.1177	-	7471B	PPB	23 Jan 2023 12:12:46	U	SMPL
O1252-01DUP - 1	MH-9DUP	495	0.0786	-	7471B	PPB	23 Jan 2023 12:15:04	U	SMPL
O1252-01MS - 1	MH-9MS	19565	3.5622	-	7471B	PPB	23 Jan 2023 12:17:21	U	SMPL
O1252-01MSD - 1	MH-9MSD	21367	3.8914	-	7471B	PPB	23 Jan 2023 12:19:36	U	SMPL
CCV97 - 1	CCV97	27200	4.9570	-	7471B	PPB	23 Jan 2023 12:21:52	U	SMPL
CCB97 - 1	CCB97	-53	-0.0215	-	7471B	PPB	23 Jan 2023 12:24:07	U	SMPL
O1232-01LX5 - 1		26	-0.0071	-	7471B	PPB	23 Jan 2023 12:26:24	U	SMPL
O1232-01A - 1		20681	3.7661	-	7471B	PPB	23 Jan 2023 12:28:40	U	SMPL
O1252-01LX5 - 1		87	0.0041	-	7471B	PPB	23 Jan 2023 12:30:55	U	SMPL
O1252-01A - 1		21214	3.8635	-	7471B	PPB	23 Jan 2023 12:33:11	U	SMPL
CCV98 - 1	CCV98	25827	4.7062	-	7471B	PPB	23 Jan 2023 12:35:27	U	SMPL
CCB98 - 1	CCB98	-53	-0.0215	-	7471B	PPB	23 Jan 2023 12:37:43	U	SMPL