

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH QC BATCHID: LB123754

	NA	NO	YES
1.Calibration Summary met criteria.	_____	_____	_____✓_____
2.ICP Interference Check Sample Results Summary Submitted.	_____	_____	_____✓_____
3. Serial Dilution Summary (if applicable) Submitted.	_____	_____	_____✓_____
4.Laboratory Control Sample Summary (if applicable) Submitted.:	_____	_____	_____✓_____
5.Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓_____	_____
6.Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
7.Sample Duplicate Analysis Met QC Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
8.Digestion Holding Time Met If not met, list number of days exceeded for each sample:	_____	_____	_____✓_____
9.Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____

ADDITIONAL COMMENTS:

MS/MSD fail for O1180-02 (Cr) due to chemical interference during digestion process. Post spike recovery passed.

O1198-01 MS/MSD fail (Sb,Cr,Co,Cu,Ni,Na,V,Zn) due to chemical interference during digestion process. Post spike recovery passed.

O1232-01 MS/MSD fail (Sb,Be,Cu) parameters due to chemical interference during digestion process. Post spike recovery passed.

Analyst bin

Date 1/24/2023 12:00:00 AM

QA REVIEW

Date