



Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Louis Berger U.S., Inc., A WSP CompanySDG No.: O1232Contract: lou01Lab Code: CHEMCase No.: O1232SAS No.: O1232ICS Source: EPAInstrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	1.24			-20	20	01/25/2023	12:33	LB123806
	Barium	11.4	6.0	190	-94	106	01/25/2023	12:33	LB123806
	Cadmium	2.75	1.0	275	-5	7	01/25/2023	12:33	LB123806
	Chromium	56.5	52.0	109	42	62	01/25/2023	12:33	LB123806
	Lead	4.10			-12	12	01/25/2023	12:33	LB123806
	Selenium	1.75			-20	20	01/25/2023	12:33	LB123806
	Silver	-0.64			-10	10	01/25/2023	12:33	LB123806
ICSAB01	Arsenic	107	100	107	85	120	01/25/2023	12:37	LB123806
	Barium	483	540	89	440	640	01/25/2023	12:37	LB123806
	Cadmium	1010	970	104	820	1100	01/25/2023	12:37	LB123806
	Chromium	565	540	105	460	620	01/25/2023	12:37	LB123806
	Lead	50.7	49.0	104	37	61	01/25/2023	12:37	LB123806
	Selenium	51.2	46.0	111	26	66	01/25/2023	12:37	LB123806
	Silver	210	200	105	170	230	01/25/2023	12:37	LB123806