



- 7 -

Client:	<u>Louis Berger U.S., Inc., A WSP Company</u>	SDG No.:	<u>01232</u>				
Contract:	loui01	Lab Code:	CHEM	Case No.:	01232	SAS No.:	01232

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB150390BS							
Mercury	ug/L	4.0	3.88		97	80 - 120	CV



Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client:

Louis Berger U.S., Inc., A WSP Company

SDG No.:

O1232

Contract:

loui01

Lab Code:

CHEM

Case No.:

O1232

SAS No.:

O1232

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB150432BS							
Arsenic	ug/L	4000	3980		100	80 - 120	P
Barium	ug/L	1000	981		98	80 - 120	P
Cadmium	ug/L	1000	971		97	80 - 120	P
Chromium	ug/L	2000	2050		102	80 - 120	P
Lead	ug/L	5000	4870		97	80 - 120	P
Selenium	ug/L	10000	10100		101	80 - 120	P
Silver	ug/L	380	379		100	80 - 120	P