

# SAMPLE DATA



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## Report of Analysis

Client:	Louis Berger U.S., Inc., A WSP Company	Date Collected:	01/18/23		
Project:	NYCDDC Phase II SCI Arthur Kill Road CEQR	Date Received:	01/19/23		
Client Sample ID:	WC-17	SDG No.:	O1232		
Lab Sample ID:	O1232-03	Matrix:	SOIL		
Analytical Method:	8015D DRO	% Solid:	91.4	Decanted:	
Sample Wt/Vol:	30.04	Units:	g	Final Vol:	1 mL
Soil Aliquot Vol:			uL	Test:	Diesel Range Organics
Extraction Type:				Injection Volume :	
GPC Factor :		PH :			
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FF011895.D	1	01/20/23 13:10	01/20/23 17:29	PB150379

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
<b>TARGETS</b>						
DRO	DRO	1920		239	1820	ug/kg
<b>SURROGATES</b>						
16416-32-3	Tetracosane-d50	10.4		37 - 130	52%	SPK: 20

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



284 Sheffield Street, Mountainside, New Jersey - 07092

Phone: (908) 789 8900 Fax: (908) 789 8922

## LAB CHRONICLE

**OrderID:** O1232  
**Client:** Louis Berger U.S., Inc., A WSP Company  
**Contact:** Jonathan Ganz

**OrderDate:** 1/19/2023 1:44:00 PM  
**Project:** NYCDDC Phase II SCI Arthur Kill Road CEQR  
**Location:** N11,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
O1232-03	WC-17	SOIL	Diesel Range Organics	8015D	01/18/23	01/20/23	01/20/23	01/19/23