



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB011723

Review By	yogesh	Review On	1/17/2023 1:50:45 PM				
Supervise By	Ankita	Supervise On	1/17/2023 3:06:42 PM				
SubDirectory	FB011723	HP Acquire Method	HP Processing Method FB011723				
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		PP21302,PP21485,PP21486,PP21487,PP21488,PP21489					
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		PP21303,PP21490					
Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	5 GRO STD		FB029115.D	17 Jan 2023 9:03		AJ/MA	Ok
2	10 GRO STD		FB029116.D	17 Jan 2023 9:44		AJ/MA	Ok
3	20 GRO STD		FB029117.D	17 Jan 2023 10:30		AJ/MA	Ok
4	50 GRO STD		FB029118.D	17 Jan 2023 11:29		AJ/MA	Ok
5	100 GRO STD		FB029119.D	17 Jan 2023 12:11		AJ/MA	Ok
6	FB011723GROICV		FB029120.D	17 Jan 2023 13:28		AJ/MA	Ok

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB012023

Review By	yogesh	Review On	1/20/2023 3:06:20 PM				
Supervise By	Ankita	Supervise On	1/23/2023 9:00:50 AM				
SubDirectory	FB012023	HP Acquire Method	HP Processing Method		FB011723		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		PP21302,PP21485,PP21486,PP21487,PP21488,PP21489					
CCC		PP21504,PP21505,PP21506					
Internal Standard/PEM							
ICV/I.BLK		PP21303,PP21490					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB029141.D	20 Jan 2023 10:04		AJ/MA	Ok
2	VBF0120S1		FB029142.D	20 Jan 2023 10:41		AJ/MA	Ok
3	VBF0120S2		FB029143.D	20 Jan 2023 11:18		AJ/MA	Ok
4	BSF0120S1		FB029144.D	20 Jan 2023 11:55		AJ/MA	Ok
5	O1232-03		FB029145.D	20 Jan 2023 12:32	vial-A , not purged	AJ/MA	Not Ok
6	O1232-03		FB029146.D	20 Jan 2023 13:11	vial-B , not purged	AJ/MA	Not Ok
7	O1233-10		FB029147.D	20 Jan 2023 13:48	vial-A , not purged	AJ/MA	Not Ok
8	O1233-10		FB029148.D	20 Jan 2023 14:25	vial-B , not purged	AJ/MA	Not Ok
9	O1232-03		FB029149.D	20 Jan 2023 15:02	vial-C	AJ/MA	Ok
10	O1232-03MS		FB029150.D	20 Jan 2023 15:39		AJ/MA	Ok,M
11	O1232-03MSD		FB029151.D	20 Jan 2023 16:16		AJ/MA	Ok
12	20 PPB GRO STD		FB029152.D	20 Jan 2023 16:54		AJ/MA	Ok,M
13	O1233-10		FB029153.D	20 Jan 2023 18:08	vial-C , Need 10X	AJ/MA	Dilution
14	20 PPB GRO STD		FB029154.D	20 Jan 2023 18:45		AJ/MA	Ok

M : Manual Integration