



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232

Continuing Calib Date: 01/20/2023 Initial Calibration Date(s): 01/17/2023 01/18/2023

Continuing Calib Time: 12:26 Initial Calibration Time(s): 19:14 02:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	6.37	6.36	6.26	6.46	-0.01
Aroclor-1254-2 (2)	6.59	6.58	6.48	6.68	-0.01
Aroclor-1254-3 (3)	6.96	6.95	6.85	7.05	-0.01
Aroclor-1254-4 (4)	7.24	7.24	7.14	7.34	0.00
Aroclor-1254-5 (5)	7.67	7.66	7.56	7.76	0.00
Tetrachloro-m-xylene	4.34	4.33	4.23	4.43	-0.01
Decachlorobiphenyl	10.11	10.10	10.00	10.20	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232

Continuing Calib Date: 01/20/2023 Initial Calibration Date(s): 01/17/2023 01/18/2023

Continuing Calib Time: 12:26 Initial Calibration Time(s): 19:14 02:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1254-1 (1)	5.47	5.47	5.37	5.57	0.00
Aroclor-1254-2 (2)	5.62	5.62	5.52	5.72	0.00
Aroclor-1254-3 (3)	6.03	6.02	5.92	6.12	-0.01
Aroclor-1254-4 (4)	6.25	6.25	6.15	6.35	0.00
Aroclor-1254-5 (5)	6.67	6.67	6.57	6.77	0.00
Tetrachloro-m-xylene	3.58	3.58	3.48	3.68	0.00
Decachlorobiphenyl	8.60	8.60	8.50	8.70	0.00