



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/17/2023 01/17/2023

Client Sample No.: CCAL02 Date Analyzed: 01/21/2023

Lab Sample No.: AR1660CCC500 Data File : PP054792.D Time Analyzed: 01:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.504	5.405	5.605	519.700	500.000	3.9
Aroclor-1016-2	5.527	5.427	5.627	512.830	500.000	2.6
Aroclor-1016-3	5.589	5.490	5.690	508.530	500.000	1.7
Aroclor-1016-4	5.688	5.588	5.788	516.130	500.000	3.2
Aroclor-1016-5	5.984	5.885	6.085	515.860	500.000	3.2
Aroclor-1260-1	7.116	7.017	7.217	531.020	500.000	6.2
Aroclor-1260-2	7.374	7.275	7.475	501.820	500.000	0.4
Aroclor-1260-3	7.735	7.637	7.837	501.240	500.000	0.2
Aroclor-1260-4	7.961	7.863	8.063	499.040	500.000	-0.2
Aroclor-1260-5	8.278	8.181	8.381	485.150	500.000	-3.0
Decachlorobiphenyl	10.092	9.994	10.194	46.080	50.000	-7.8
Tetrachloro-m-xylene	4.333	4.231	4.431	50.890	50.000	1.8



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/17/2023 01/17/2023

Client Sample No.: CCAL02 Date Analyzed: 01/21/2023

Lab Sample No.: AR1660CCC500 Data File : PP054792.D Time Analyzed: 01:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.665	4.565	4.765	489.800	500.000	-2.0
Aroclor-1016-2	4.684	4.584	4.784	491.850	500.000	-1.6
Aroclor-1016-3	4.860	4.761	4.961	487.940	500.000	-2.4
Aroclor-1016-4	4.902	4.804	5.004	463.880	500.000	-7.2
Aroclor-1016-5	5.116	5.017	5.217	492.150	500.000	-1.6
Aroclor-1260-1	6.153	6.055	6.255	480.180	500.000	-4.0
Aroclor-1260-2	6.342	6.243	6.443	498.600	500.000	-0.3
Aroclor-1260-3	6.496	6.398	6.598	491.130	500.000	-1.8
Aroclor-1260-4	6.970	6.872	7.072	492.710	500.000	-1.5
Aroclor-1260-5	7.213	7.115	7.315	496.280	500.000	-0.7
Decachlorobiphenyl	8.598	8.502	8.702	47.510	50.000	-5.0
Tetrachloro-m-xylene	3.579	3.478	3.678	47.200	50.000	-5.6