

Project:

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Date Collected:

Date Received:

SDG No.:

Final Vol:

Test:

Matrix: % Solid:

01/18/23

01/19/23

O1232

**SOIL** 

91.4

10000

PCB

Decanted:

иL

## **Report of Analysis**

Client: Louis Berger U.S., Inc., A WSP Company

NYCDDC Phase II SCI Arthur Kill Road CEQR

Client Sample ID: WC-17

Lab Sample ID: O1232-03

Analytical Method: SW8082A

Sample Wt/Vol: 30.03 Units: g

Soil Aliquot Vol: uL

Extraction Type: Injection Volume :

GPC Factor: 1.0 PH:

Prep Method: SW3541B

 File ID/Qc Batch:
 Dilution:
 Prep Date
 Date Analyzed
 Prep Batch ID

 PP054778.D
 1
 01/20/23 10:35
 01/20/23 20:54
 PB150372

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.30	U	3.30	18.6	ug/kg
11104-28-2	Aroclor-1221	5.10	U	5.10	18.6	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	18.6	ug/kg
53469-21-9	Aroclor-1242	42.6	P	2.60	18.6	ug/kg
12672-29-6	Aroclor-1248	3.30	U	3.30	18.6	ug/kg
11097-69-1	Aroclor-1254	4.60	U	4.60	18.6	ug/kg
37324-23-5	Aroclor-1262	3.60	U	3.60	18.6	ug/kg
11100-14-4	Aroclor-1268	6.30	U	6.30	18.6	ug/kg
11096-82-5	Aroclor-1260	3.50	U	3.50	18.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.8		40 - 162	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.0		32 - 176	105%	SPK: 20

## Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit