

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **Report of Analysis**

Client: Louis Berger U.S., Inc., A WSP Company Date Collected:

01/18/23

PCB

Project:

NYCDDC Phase II SCI Arthur Kill Road CEQR

Date Received: 01/19/23

Client Sample ID:

B-P18A

SDG No.: O1232

Lab Sample ID:

O1232-04

**SOIL** 

Analytical Method:

SW8082A

Matrix:

87.5

Sample Wt/Vol:

% Solid:

Decanted:

30.06 Units: Final Vol:

10000 иL

Soil Aliquot Vol:

g uL

PH:

Test:

Extraction Type:

1.0

GPC Factor:

Injection Volume:

Prep Method:

File ID/Qc Batch:

SW3541B

Dilution:

Prep Date

Date Analyzed

Prep Batch ID

PP054779.D

1

01/20/23 10:35

01/20/23 21:11

PB150372

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.50	U	3.50	19.4	ug/kg
11104-28-2	Aroclor-1221	5.40	U	5.40	19.4	ug/kg
11141-16-5	Aroclor-1232	4.50	U	4.50	19.4	ug/kg
53469-21-9	Aroclor-1242	2.80	U	2.80	19.4	ug/kg
12672-29-6	Aroclor-1248	3.40	U	3.40	19.4	ug/kg
11097-69-1	Aroclor-1254	4.80	U	4.80	19.4	ug/kg
37324-23-5	Aroclor-1262	3.80	U	3.80	19.4	ug/kg
11100-14-4	Aroclor-1268	6.50	U	6.50	19.4	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.6		40 - 162	113%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.2		32 - 176	106%	SPK: 20

## Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit