

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **Report of Analysis**

Client: Louis Berger U.S., Inc., A WSP Company Date Collected: Date Received:

Matrix:

Project:

NYCDDC Phase II SCI Arthur Kill Road CEQR

SDG No.:

Client Sample ID:

PB150372BL

O1232

Lab Sample ID:

PB150372BL

10000

Analytical Method:

SW8082A

**SOIL** 

Sample Wt/Vol:

Units: 30

% Solid: 100

Final Vol:

Decanted: иL

Soil Aliquot Vol:

иL

Test: PCB

Extraction Type:

GPC Factor:

1.0

PH:

Injection Volume:

Prep Method:

File ID/Qc Batch:

SW3541B

Dilution:

Date Analyzed

Prep Batch ID

PP054772.D

1

01/20/23 10:35

Prep Date

01/20/23 19:15

PB150372

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TADCETS						
TARGETS 12674-11-2	Aroclor-1016	3.10	U	3.10	17.0	ug/kg
			_			
11104-28-2	Aroclor-1221	4.70	U	4.70	17.0	ug/kg
11141-16-5	Aroclor-1232	3.90	U	3.90	17.0	ug/kg
53469-21-9	Aroclor-1242	2.40	U	2.40	17.0	ug/kg
12672-29-6	Aroclor-1248	3.00	U	3.00	17.0	ug/kg
11097-69-1	Aroclor-1254	4.20	U	4.20	17.0	ug/kg
37324-23-5	Aroclor-1262	3.30	U	3.30	17.0	ug/kg
11100-14-4	Aroclor-1268	5.70	U	5.70	17.0	ug/kg
11096-82-5	Aroclor-1260	3.20	U	3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.6		40 - 162	83%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.6		32 - 176	88%	SPK: 20

## Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit