



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: loui01

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232

Instrument ID: ECD_L Date(s) Analyzed: 01/20/2023 01/20/2023

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.03	5.93	6.13	16579700
		2	6.84	6.74	6.94	67395000
		3	6.93	6.83	7.03	48523600
		4	7.03	6.93	7.13	26898900
		5	7.35	7.25	7.45	37572200



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: loui01

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232

Instrument ID: ECD_L Date(s) Analyzed: 01/20/2023 01/20/2023

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.24	5.14	5.34	13842900
		2	5.88	5.78	5.98	16444600
		3	6.52	6.42	6.62	53738800
		4	6.65	6.55	6.75	73380200
		5	6.96	6.86	7.06	31164000