

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/20/2023 01/20/2023

Client Sample No.: CCAL03 Date Analyzed: 01/21/2023

Lab Sample No.: PSTDCCC050 Data File : PL080454.D Time Analyzed: 03:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.501	6.404	6.604	50.450	50.000	0.9
4,4'-DDE	5.987	5.889	6.089	50.250	50.000	0.5
4,4'-DDT	6.814	6.716	6.916	48.260	50.000	-3.5
Aldrin	5.050	4.952	5.152	50.390	50.000	0.8
alpha-BHC	3.799	3.701	3.901	51.560	50.000	3.1
alpha-Chlordane	5.809	5.711	5.911	50.040	50.000	0.1
beta-BHC	4.324	4.225	4.425	49.280	50.000	-1.4
Decachlorobiphenyl	8.789	8.693	8.893	48.200	50.000	-3.6
delta-BHC	4.567	4.468	4.668	51.820	50.000	3.6
Dieldrin	6.131	6.033	6.233	50.070	50.000	0.1
Endosulfan I	5.857	5.759	5.959	50.130	50.000	0.3
Endosulfan II	6.577	6.480	6.680	43.040	50.000	-13.9
Endosulfan sulfate	6.940	6.843	7.043	50.000	50.000	0.0
Endrin	6.358	6.261	6.461	53.860	50.000	7.7
Endrin aldehyde	6.706	6.609	6.809	48.530	50.000	-2.9
Endrin ketone	7.421	7.324	7.524	49.890	50.000	-0.2
gamma-BHC (Lindane)	4.128	4.029	4.229	51.830	50.000	3.7
gamma-Chlordane	5.728	5.632	5.832	50.570	50.000	1.1
Heptachlor	4.712	4.613	4.813	50.890	50.000	1.8
Heptachlor epoxide	5.475	5.377	5.577	51.910	50.000	3.8
Methoxychlor	7.292	7.194	7.394	48.720	50.000	-2.6
Tetrachloro-m-xylene	3.351	3.252	3.452	50.200	50.000	0.4



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/20/2023 01/20/2023

Client Sample No.: CCAL03 Date Analyzed: 01/21/2023

Lab Sample No.: PSTDCCC050 Data File : PL080454.D Time Analyzed: 03:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.703	5.606	5.806	51.750	50.000	3.5
4,4'-DDE	5.150	5.052	5.252	50.440	50.000	0.9
4,4'-DDT	5.955	5.858	6.058	50.090	50.000	0.2
Aldrin	4.139	4.041	4.241	52.640	50.000	5.3
alpha-BHC	3.190	3.091	3.291	51.070	50.000	2.1
alpha-Chlordane	4.957	4.859	5.059	48.930	50.000	-2.1
beta-BHC	3.815	3.716	3.916	49.760	50.000	-0.5
Decachlorobiphenyl	7.838	7.742	7.942	48.250	50.000	-3.5
delta-BHC	4.045	3.946	4.146	51.970	50.000	3.9
Dieldrin	5.276	5.179	5.379	50.620	50.000	1.2
Endosulfan I	5.011	4.913	5.113	47.620	50.000	-4.8
Endosulfan II	5.844	5.748	5.948	52.940	50.000	5.9
Endosulfan sulfate	6.248	6.151	6.351	49.390	50.000	-1.2
Endrin	5.551	5.454	5.654	50.690	50.000	1.4
Endrin aldehyde	6.026	5.928	6.128	47.950	50.000	-4.1
Endrin ketone	6.753	6.656	6.856	49.690	50.000	-0.6
gamma-BHC (Lindane)	3.519	3.420	3.620	50.960	50.000	1.9
gamma-Chlordane	4.893	4.795	4.995	49.060	50.000	-1.9
Heptachlor	3.860	3.762	3.962	49.840	50.000	-0.3
Heptachlor epoxide	4.642	4.544	4.744	49.640	50.000	-0.7
Methoxychlor	6.534	6.436	6.636	48.790	50.000	-2.4
Tetrachloro-m-xylene	2.689	2.590	2.790	49.770	50.000	-0.5