



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: O1232

Client: Louis Berger U.S., Inc., A WSP Compan

Analytical Method: 8081B

Datafile : PL080430.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB150373BS	alpha-BHC	16.66	15.1	ug/kg	91				84	123	
	beta-BHC	16.66	14.3	ug/kg	86				82	123	
	delta-BHC	16.66	15.2	ug/kg	91				83	126	
	gamma-BHC (Lindane)	16.66	15.1	ug/kg	91				83	125	
	Heptachlor	16.66	15.2	ug/kg	91				83	122	
	Aldrin	16.66	15.5	ug/kg	93				82	124	
	Heptachlor epoxide	16.66	15.3	ug/kg	92				83	120	
	Endosulfan I	16.66	15.0	ug/kg	90				81	124	
	Dieldrin	16.66	15.9	ug/kg	95				85	121	
	4,4'-DDE	16.66	15.3	ug/kg	92				81	123	
	Endrin	16.66	16.5	ug/kg	99				76	130	
	Endosulfan II	16.66	17.4	ug/kg	104				80	125	
	4,4'-DDD	16.66	15.2	ug/kg	91				80	131	
	Endosulfan sulfate	16.66	15.6	ug/kg	94				81	122	
	4,4'-DDT	16.66	15.0	ug/kg	90				70	129	
	Methoxychlor	16.66	15.5	ug/kg	93				78	129	
	Endrin ketone	16.66	15.5	ug/kg	93				77	132	
	Endrin aldehyde	16.66	16.4	ug/kg	98				79	124	
	alpha-Chlordane	16.66	15.3	ug/kg	92				84	120	
	gamma-Chlordane	16.66	15.0	ug/kg	90				83	122	