



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG No.: O1232

Instrument ID: BNA_F Calibration Date/Time: 01/20/2023 16:40

Lab File ID: BF132164.D Init. Calib. Date(s): 01/19/2023 01/19/2023

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:21 15:11

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.145	1.115		-2.6	
Phenol-d6	1.441	1.412		-2.0	
Nitrobenzene-d5	0.376	0.393		4.5	
Naphthalene	1.027	1.008		-1.9	
2-Fluorobiphenyl	1.369	1.224		-10.6	
Acenaphthylene	1.951	1.915		-1.8	
Acenaphthene	1.174	1.145		-2.5	20.0
Fluorene	1.344	1.292		-3.9	
2,4,6-Tribromophenol	0.214	0.231		7.9	
Phenanthrene	1.050	1.000		-4.8	
Anthracene	1.062	1.012		-4.7	
Fluoranthene	1.078	1.118		3.7	20.0
Pyrene	1.843	1.942		5.4	
Terphenyl-d14	1.194	1.203		0.8	
Benzo (a) anthracene	1.429	1.449		1.4	
Chrysene	1.360	1.334		-1.9	
Benzo (b) fluoranthene	1.218	1.276		4.8	
Benzo (k) fluoranthene	1.242	1.197		-3.6	
Benzo (a) pyrene	1.171	1.185		1.2	20.0
Indeno (1,2,3-cd) pyrene	1.451	1.547		6.6	
Dibenzo (a,h) anthracene	1.184	1.248		5.4	
Benzo (g,h,i) perylene	1.246	1.338		7.4	

All other compounds must meet a minimum RRF of 0.010.