



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG No.: O1232

Instrument ID: BNA_P Calibration Date/Time: 01/24/2023 15:13

Lab File ID: BP013571.D Init. Calib. Date(s): 01/23/2023 01/23/2023

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:25 16:01

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.167	1.147		-1.7	
Phenol-d6	1.571	1.597		1.7	
Nitrobenzene-d5	0.303	0.333		9.9	
Naphthalene	1.075	1.033		-3.9	
2-Fluorobiphenyl	1.457	1.331		-8.6	
Acenaphthylene	1.968	1.875		-4.7	
Acenaphthene	1.216	1.183		-2.7	20.0
Fluorene	1.482	1.443		-2.6	
2,4,6-Tribromophenol	0.201	0.241		19.9	
Phenanthrene	1.113	1.061		-4.7	
Anthracene	1.129	1.078		-4.5	
Fluoranthene	1.183	1.251		5.7	20.0
Pyrene	1.529	1.345		-12.0	
Terphenyl-d14	1.112	1.000		-10.1	
Benzo (a) anthracene	1.393	1.340		-3.8	
Chrysene	1.333	1.268		-4.9	
Benzo (b) fluoranthene	1.230	1.204		-2.1	
Benzo (k) fluoranthene	1.255	1.230		-2.0	
Benzo (a) pyrene	1.194	1.155		-3.3	20.0
Indeno (1,2,3-cd) pyrene	1.473	1.333		-9.5	
Dibenzo (a,h) anthracene	1.236	1.129		-8.7	
Benzo (g,h,i) perylene	1.206	1.048		-13.1	

All other compounds must meet a minimum RRF of 0.010.