



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary
SW-846

SDG No.: O1232

Client: Louis Berger U.S., Inc., A WSP Company

Analytical Method: SW8270E

Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
		Result	Result			Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	O1233-02MSD	Client Sample ID:	B-P34AMSD					DataFile:	BF132172.D		
Naphthalene	1900	0	1600	ug/Kg	84		5		72	110	20
Acenaphthylene	1900	0	1600	ug/Kg	84		5		79	118	20
Acenaphthene	1900	0	1800	ug/Kg	95		5		70	121	20
Fluorene	1900	0	1600	ug/Kg	84		5		68	116	20
Phenanthrene	1900	0	1500	ug/Kg	79		5		72	113	20
Anthracene	1900	0	1500	ug/Kg	79		5		62	124	20
Fluoranthene	1900	0	1700	ug/Kg	89		5		59	125	20
Pyrene	1900	0	1600	ug/Kg	84		5		52	128	20
Benzo(a)anthracene	1900	0	1600	ug/Kg	84		1		71	114	20
Chrysene	1900	0	1600	ug/Kg	84		5		57	121	20
Benzo(b)fluoranthene	1900	0	1800	ug/Kg	95		5		67	121	20
Benzo(k)fluoranthene	1900	0	1700	ug/Kg	89		5		74	114	20
Benzo(a)pyrene	1900	0	1600	ug/Kg	84		5		70	142	20
Indeno(1,2,3-cd)pyrene	1900	0	1500	ug/Kg	79		5		61	125	20
Dibenz(a,h)anthracene	1900	0	1600	ug/Kg	84		1		67	130	20
Benzo(g,h,i)perylene	1900	0	1500	ug/Kg	79		5		53	140	20