



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: loui01

Lab Code: CHEM

SAS No.: 01232

SDG NO.: 01232

Lab File ID: BP013544.D

DFTPP Injection Date: 01/23/2023

Instrument ID: BNA_P

DFTPP Injection Time: 10:51

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	38.4
68	Less than 2.0% of mass 69	0.6 (1.7) 1
69	Mass 69 relative abundance	34.2
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	47.0
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	26.6
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.9 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP013545.D	01/23/2023	11:25
SSTDICC005	SSTDICC005	BP013546.D	01/23/2023	12:00
SSTDICC010	SSTDICC010	BP013547.D	01/23/2023	12:34
SSTDICC020	SSTDICC020	BP013548.D	01/23/2023	13:08
SSTDICCC040	SSTDICCC040	BP013549.D	01/23/2023	13:42
SSTDICC050	SSTDICC050	BP013550.D	01/23/2023	14:17
SSTDICC060	SSTDICC060	BP013551.D	01/23/2023	14:51
SSTDICC080	SSTDICC080	BP013552.D	01/23/2023	16:01