



SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01
 Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG No.: O1232
 Instrument ID: BNA_P Calibration Date/Time: 01/24/2023 15:13
 Lab File ID: BP013571.D Init. Calib. Date(s): 01/23/2023 01/23/2023
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:25 16:01
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.167	1.147		-1.7	
Benzaldehyde	0.995	0.888		-10.8	
Phenol-d6	1.571	1.597		1.7	
Phenol	1.628	1.651		1.4	20.0
bis(2-Chloroethyl)ether	1.331	1.329		-0.2	
2-Chlorophenol	1.288	1.323		2.7	
2-Methylphenol	1.182	1.208		2.2	
2,2-oxybis(1-Chloropropane)	1.822	1.794		-1.5	
Acetophenone	0.513	0.490		-4.5	
3+4-Methylphenols	1.556	1.656		6.4	
n-Nitroso-di-n-propylamine	1.012	1.063	0.050	5.0	
Nitrobenzene-d5	0.303	0.333		9.9	
Hexachloroethane	0.483	0.491		1.7	
Nitrobenzene	0.326	0.344		5.5	
Isophorone	0.719	0.697		-3.1	
2-Nitrophenol	0.094	0.142		51.1	20.0
2,4-Dimethylphenol	0.310	0.311		0.3	
bis(2-Chloroethoxy)methane	0.433	0.428		-1.2	
2,4-Dichlorophenol	0.285	0.299		4.9	20.0
Naphthalene	1.075	1.033		-3.9	
4-Chloroaniline	0.476	0.475		-0.2	
Hexachlorobutadiene	0.189	0.179		-5.3	20.0
Caprolactam	0.093	0.109		17.2	
4-Chloro-3-methylphenol	0.311	0.339		9.0	20.0
2-Methylnaphthalene	0.764	0.764		0.0	
Hexachlorocyclopentadiene	0.290	0.295	0.050	1.7	
2,4,6-Trichlorophenol	0.360	0.382		6.1	20.0
2-Fluorobiphenyl	1.457	1.331		-8.6	
2,4,5-Trichlorophenol	0.426	0.453		6.3	
1,1-Biphenyl	1.606	1.480		-7.8	
2-Chloronaphthalene	1.203	1.104		-8.2	
2-Nitroaniline	0.241	0.326		35.3	
Dimethylphthalate	1.487	1.504		1.1	
Acenaphthylene	1.968	1.875		-4.7	
2,6-Dinitrotoluene	0.240	0.313		30.4	
3-Nitroaniline	0.275	0.361		31.3	
Acenaphthene	1.216	1.183		-2.7	20.0
2,4-Dinitrophenol	0.058	0.118	0.050	103.4	
4-Nitrophenol	0.212	0.288	0.050	35.8	
Dibenzofuran	1.864	1.797		-3.6	
2,4-Dinitrotoluene	0.266	0.421		58.3	



SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01
 Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG No.: O1232
 Instrument ID: BNA_P Calibration Date/Time: 01/24/2023 15:13
 Lab File ID: BP013571.D Init. Calib. Date(s): 01/23/2023 01/23/2023
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:25 16:01
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Diethylphthalate	1.452	1.521		4.8	
4-Chlorophenyl-phenylether	0.729	0.712		-2.3	
Fluorene	1.482	1.443		-2.6	
4-Nitroaniline	0.265	0.376		41.9	
4,6-Dinitro-2-methylphenol	0.058	0.084		44.8	
n-Nitrosodiphenylamine	0.652	0.591		-9.4	20.0
2,4,6-Tribromophenol	0.201	0.241		19.9	
4-Bromophenyl-phenylether	0.222	0.200		-9.9	
Hexachlorobenzene	0.239	0.222		-7.1	
Atrazine	0.187	0.197		5.3	
Pentachlorophenol	0.139	0.160		15.1	20.0
Phenanthrene	1.113	1.061		-4.7	
Anthracene	1.129	1.078		-4.5	
Carbazole	1.015	1.015		0.0	
Di-n-butylphthalate	1.130	1.228		8.7	
Fluoranthene	1.183	1.251		5.7	20.0
Pyrene	1.529	1.345		-12.0	
Terphenyl-d14	1.112	1.000		-10.1	
Butylbenzylphthalate	0.487	0.567		16.4	
3,3-Dichlorobenzidine	0.449	0.464		3.3	
Benzo(a)anthracene	1.393	1.340		-3.8	
Chrysene	1.333	1.268		-4.9	
Bis(2-ethylhexyl)phthalate	0.779	0.839		7.7	
Di-n-octyl phthalate	1.256	1.423		13.3	20.0
Benzo(b)fluoranthene	1.230	1.204		-2.1	
Benzo(k)fluoranthene	1.255	1.230		-2.0	
Benzo(a)pyrene	1.194	1.155		-3.3	20.0
Indeno(1,2,3-cd)pyrene	1.473	1.333		-9.5	
Dibenzo(a,h)anthracene	1.236	1.129		-8.7	
Benzo(g,h,i)perylene	1.206	1.048		-13.1	
1,2,4,5-Tetrachlorobenzene	0.616	0.561		-8.9	
1,4-Dioxane	0.471	0.449		-4.7	20.0
2,3,4,6-Tetrachlorophenol	0.327	0.373		14.1	

All other compounds must meet a minimum RRF of 0.010.