



NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bg012423

SequenceID : bg012423

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	<u>✓</u>	<u> </u>	<u> </u>
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	<u>✓</u>	<u> </u>	<u> </u>
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	<u>✓</u>	<u> </u>	<u> </u>
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	<u>✓</u>	<u> </u>	<u> </u>
5. GC/MS Calibration Met:	<u>✓</u>	<u> </u>	<u> </u>
a. Initial calibration Meet Criteria	<u>✓</u>	<u> </u>	<u> </u>
If not met, list those compounds and their recoveries which fall outside the acceptable range.	<u> </u>	<u> </u>	<u> </u>
b. Continuous Calibration(CCC) Meet Criteria	<u> </u>	<u>✓</u>	<u> </u>
If not met, list those compounds and their recoveries which fall outside the acceptable range.	<u> </u>	<u> </u>	<u> </u>
In the SSTDCCC[BG056397.D], compound #16 is biased high but not present and compound #32 is marginally biased low but not required by any of the analyzed samples.			
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	<u> </u>	<u>✓</u>	<u> </u>
a. B/N Fraction			
d. Acid Fraction			

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

____ ✓

a. B/N Fraction

d. Acid Fraction

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

____ ✓

a. B/N Fraction The recovery of one compound failed in O1260-22MS/MSD due to matrix interference and no corrective action is required.

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

____ ✓

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

____ ✓

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

____ ✓

ADDITIONAL COMMENTS:

nishadh

Analyst

01/25/2023

Date



NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bg012423

SequenceID : bg012423 NA NO YES