



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH
Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232
EPA Sample No.: SSTDCCC040 Date Analyzed: 01/20/2023
Lab File ID: BF132164.D Time Analyzed: 16:40
Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	175584	7.09	590884	8.38	314145	10.14
UPPER LIMIT	351168	7.59	1181770	8.878	628290	10.642
LOWER LIMIT	87792	6.59	295442	7.878	157073	9.642
EPA SAMPLE NO.						
01 B-P-17A	179173	7.09	654165	8.38	350949	10.14
02 B-P-35A	185310	7.09	674589	8.38	371842	10.14
03 B-P34AMS	184212	7.09	674454	8.38	369043	10.14
04 B-P34AMSD	185010	7.09	669898	8.38	363370	10.14

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.