



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232

EPA Sample No.: SSTDCCC040 Date Analyzed: 01/24/2023

Lab File ID: BP013562.D Time Analyzed: 09:53

Instrument ID: BNA\_P GC Column: ZB-GR ID: 0.25 (mm)

|                | IS1 (DCB)<br>AREA # | RT #  | IS2 (NPT)<br>AREA # | RT #   | IS3 (ANT)<br>AREA # | RT #   |
|----------------|---------------------|-------|---------------------|--------|---------------------|--------|
| 12 HOUR STD    | 306826              | 8.164 | 1252680             | 10.99  | 763071              | 14.80  |
| UPPER LIMIT    | 613652              | 8.664 | 2505360             | 11.493 | 1526140             | 15.299 |
| LOWER LIMIT    | 153413              | 7.664 | 626340              | 10.493 | 381536              | 14.299 |
| EPA SAMPLE NO. |                     |       |                     |        |                     |        |
| 01 B-P-17B     | 250560              | 8.16  | 955341              | 10.99  | 503514              | 14.79  |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.