



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232

EPA Sample No.: SSTDCCC040 Date Analyzed: 01/20/2023

Lab File ID: BF132164.D Time Analyzed: 16:40

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	631542	11.642	364265	14.307	316916	15.901
UPPER LIMIT	1263080	12.142	728530	14.807	633832	16.401
LOWER LIMIT	315771	11.142	182133	13.807	158458	15.401
EPA SAMPLE NO.						
01 B-P-17A	681798	11.64	457343	14.30	397282	15.90
02 B-P-35A	723544	11.64	486031	14.30	396449	15.90
03 B-P34AMS	748390	11.64	455426	14.31	405760	15.90
04 B-P34AMSD	736982	11.64	439968	14.31	406965	15.90

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.