



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232

EPA Sample No.: SSTDCCC040

Date Analyzed: 01/24/2023

Lab File ID: BG056403.D

Time Analyzed: 14:59

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	289164	17.648	265057	21.943	286788	25.369
UPPER LIMIT	578328	18.148	530114	22.443	573576	25.869
LOWER LIMIT	144582	17.148	132529	21.443	143394	24.869
EPA SAMPLE NO.						
01 B-P-19B	291766	17.64	274867	21.93	299958	25.36

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.