



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: 01232 SAS No.: 01232 SDG NO.: 01232

EPA Sample No.: SSTDCCC040 Date Analyzed: 01/24/2023

Lab File ID: BP013562.D Time Analyzed: 09:53

Instrument ID: BNA\_P GC Column: ZB-GR ID: 0.25 (mm)

|                | IS4 (PHN)<br>AREA # | RT #   | IS5 (CRY)<br>AREA # | RT #   | IS6 (PRY)<br>AREA # | RT #   |
|----------------|---------------------|--------|---------------------|--------|---------------------|--------|
| 12 HOUR STD    | 1590580             | 17.551 | 1270420             | 21.628 | 1120190             | 24.216 |
| UPPER LIMIT    | 3181160             | 18.051 | 2540840             | 22.128 | 2240380             | 24.716 |
| LOWER LIMIT    | 795290              | 17.051 | 635210              | 21.128 | 560095              | 23.716 |
| EPA SAMPLE NO. |                     |        |                     |        |                     |        |
| 01 B-P-17B     | 1009560             | 17.55  | 942607              | 21.62  | 1148370             | 24.21  |

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.