



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: loui01

Lab Code: CHEM

SAS No.: O1232

SDG NO.: O1232

Lab File ID: BP013561.D

DFTPP Injection Date: 01/24/2023

Instrument ID: BNA\_P

DFTPP Injection Time: 09:20

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.3
68	Less than 2.0% of mass 69	0.4 ( 1.6 ) 1
69	Mass 69 relative abundance	30.6
70	Less than 2.0% of mass 69	0.1 ( 0.5 ) 1
127	10.0 - 80.0% of mass 198	44.6
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	27.3
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	17.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.5 ( 19.5 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP013562.D	01/24/2023	09:53
B-P-17B	O1232-02	BP013566.D	01/24/2023	12:17