



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: loui01

Lab Code: CHEM

SAS No.: O1232

SDG NO.: O1232

Lab File ID: BP013570.D

DFTPP Injection Date: 01/24/2023

Instrument ID: BNA_P

DFTPP Injection Time: 14:38

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	34.9
68	Less than 2.0% of mass 69	0.5 (1.7) 1
69	Mass 69 relative abundance	32.2
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	46.1
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	26.4
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	15.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.5 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP013571.D	01/24/2023	15:13
PB150378BL	PB150378BL	BP013572.D	01/24/2023	15:50
B-P-35B	O1232-09	BP013581.D	01/24/2023	21:02
B-P-18A	O1232-04	BP013586.D	01/24/2023	23:52