



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232

EPA Sample No.: SSTDCCC040 Date Analyzed: 01/20/2023

Lab File ID: BG056363.D Time Analyzed: 11:44

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	37412	8.307	143761	11.14	115862	14.92
UPPER LIMIT	74824	8.807	287522	11.639	231724	15.423
LOWER LIMIT	18706	7.807	71880.5	10.639	57931	14.423
EPA SAMPLE NO.						
01 B-P18B	44014	8.30	169116	11.13	128743	14.92
02 B-P19A	36892	8.30	139086	11.13	114800	14.91
03 DUP01	47011	8.30	177848	11.13	127856	14.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.