



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH
Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232
EPA Sample No.: SSTDCCC040 Date Analyzed: 01/24/2023
Lab File ID: BG056403.D Time Analyzed: 14:59
Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	39529	8.294	145808	11.13	117421	14.91
UPPER LIMIT	79058	8.794	291616	11.626	234842	15.41
LOWER LIMIT	19764.5	7.794	72904	10.626	58710.5	14.41
EPA SAMPLE NO.						
01 B-P19B	40835	8.29	164752	11.13	126513	14.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.