

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232

EPA Sample No.: SSTDCCC040 Date Analyzed: 01/24/2023

Lab File ID: BP013571.D Time Analyzed: 15:13

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1830850	17.551	1775330	21.627	1903850	24.215
UPPER LIMIT	3661700	18.051	3550660	22.127	3807700	24.715
LOWER LIMIT	915425	17.051	887665	21.127	951925	23.715
EPA SAMPLE NO.						
01 PB150378BL	1910080	17.55	1883140	21.63	2131870	24.22
02 B-P18A	2094750	17.55	1475320	21.62	1445620	24.20
03 B-P35B	1874320	17.55	1653010	21.62	1474510	24.20

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.