



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: loui01

Lab Code: CHEM

SAS No.: 01232 SDG NO.: 01232

Lab File ID: BF132139.D

DFTPP Injection Date: 01/19/2023

Instrument ID: BNA_F

DFTPP Injection Time: 10:46

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	40.6
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	37.4
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	48.4
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	28.2
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	12.5
442	Greater than 50% of mass 198	82.8
443	15.0 - 24.0% of mass 442	16 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF132140.D	01/19/2023	11:21
SSTDICC005	SSTDICC005	BF132141.D	01/19/2023	11:53
SSTDICC010	SSTDICC010	BF132142.D	01/19/2023	12:26
SSTDICC020	SSTDICC020	BF132143.D	01/19/2023	12:59
SSTDICCC040	SSTDICCC040	BF132144.D	01/19/2023	13:32
SSTDICC050	SSTDICC050	BF132145.D	01/19/2023	14:05
SSTDICC060	SSTDICC060	BF132146.D	01/19/2023	14:38
SSTDICC080	SSTDICC080	BF132147.D	01/19/2023	15:11