



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: loui01

Lab Code: CHEM

SAS No.: 01232 SDG NO.: 01232

Lab File ID: BF132163.D

DFTPP Injection Date: 01/20/2023

Instrument ID: BNA_F

DFTPP Injection Time: 16:08

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	38.5
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	36
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	46.9
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	28.3
365	Greater than 1% of mass 198	4.3
441	Present, but less than mass 443	13.9
442	Greater than 50% of mass 198	89.3
443	15.0 - 24.0% of mass 442	17.6 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF132164.D	01/20/2023	16:40
B-P34AMS	O1233-02MS	BF132171.D	01/20/2023	20:35
B-P34AMSD	O1233-02MSD	BF132172.D	01/20/2023	21:08
B-P17A	O1232-01	BF132175.D	01/20/2023	22:46
B-P35A	O1232-08	BF132177.D	01/20/2023	23:51