



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: loui01

Lab Code: CHEM

SAS No.: 01232 SDG NO.: 01232

Lab File ID: BP013668.D

DFTPP Injection Date: 01/31/2023

Instrument ID: BNA_P

DFTPP Injection Time: 10:47

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	24
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	27.1
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	34.4
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.3
275	10.0 - 60.0% of mass 198	27.4
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	15.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.5 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP013669.D	01/31/2023	11:21
SSTDICC005	SSTDICC005	BP013670.D	01/31/2023	11:55
SSTDICC010	SSTDICC010	BP013671.D	01/31/2023	12:30
SSTDICC020	SSTDICC020	BP013672.D	01/31/2023	13:04
SSTDICCC040	SSTDICCC040	BP013673.D	01/31/2023	13:38
SSTDICC050	SSTDICC050	BP013674.D	01/31/2023	14:12
SSTDICC060	SSTDICC060	BP013675.D	01/31/2023	14:47
SSTDICC080	SSTDICC080	BP013676.D	01/31/2023	15:56
PB150378BS	PB150378BS	BP013684.D	01/31/2023	21:04