

Instrument ID: MSVOA\_Y

### Daily Analysis Runlog For Sequence/QC Batch ID # VY011623

|                                         |               |                                                                   |                      |                      |              |  |  |
|-----------------------------------------|---------------|-------------------------------------------------------------------|----------------------|----------------------|--------------|--|--|
| Review By                               | Krupa Patel   | Review On                                                         | 1/17/2023 2:36:16 PM |                      |              |  |  |
| Supervise By                            | Mahesh Dadoda | Supervise On                                                      | 1/17/2023 2:38:16 PM |                      |              |  |  |
| SubDirectory                            | VY011623      | HP Acquire Method                                                 | MSVOA_Y              | HP Processing Method | 82y011623s.m |  |  |
| STD. NAME                               |               | STD REF.#                                                         |                      |                      |              |  |  |
| Tune/Reschk<br>Initial Calibration Stds |               | VP117988<br>VP117989,VP117990,VP117991,VP117992,VP117993,VP117994 |                      |                      |              |  |  |
| CCC                                     |               | VP117996                                                          |                      |                      |              |  |  |
| Internal Standard/PEM                   |               | VP116738                                                          |                      |                      |              |  |  |
| ICV/I.BLK                               |               | VP117995                                                          |                      |                      |              |  |  |
| Surrogate Standard                      |               |                                                                   |                      |                      |              |  |  |
| MS/MSD Standard                         |               |                                                                   |                      |                      |              |  |  |
| LCS Standard                            |               |                                                                   |                      |                      |              |  |  |

| Sr# | SampleId     | ClientID     | Data File Name | Date-Time         | Comment                 | Operator | Status   |
|-----|--------------|--------------|----------------|-------------------|-------------------------|----------|----------|
| 1   | BFB          | BFB          | VY012193.D     | 16 Jan 2023 09:00 |                         | KP/MD    | Ok       |
| 2   | VSTDICC005   | VSTDICC005   | VY012194.D     | 16 Jan 2023 12:17 |                         | KP/MD    | Ok,M     |
| 3   | VSTDICC010   | VSTDICC010   | VY012195.D     | 16 Jan 2023 13:03 |                         | KP/MD    | Ok,M     |
| 4   | VSTDICC020   | VSTDICC020   | VY012196.D     | 16 Jan 2023 13:26 |                         | KP/MD    | Ok,M     |
| 5   | VSTDICCC050  | VSTDICCC050  | VY012197.D     | 16 Jan 2023 13:48 |                         | KP/MD    | Ok,M     |
| 6   | VSTDICC100   | VSTDICC100   | VY012198.D     | 16 Jan 2023 14:11 |                         | KP/MD    | Ok,M     |
| 7   | VSTDICC150   | VSTDICC150   | VY012199.D     | 16 Jan 2023 14:34 |                         | KP/MD    | Ok,M     |
| 8   | VIBLK        | VIBLK        | VY012200.D     | 16 Jan 2023 14:58 |                         | KP/MD    | Ok       |
| 9   | VSTDICV050   | ICVVY011623  | VY012201.D     | 16 Jan 2023 15:21 |                         | KP/MD    | Ok,M     |
| 10  | VY0116SBL01  | VY0116SBL01  | VY012202.D     | 16 Jan 2023 16:21 |                         | KP/MD    | Ok       |
| 11  | O1100-05     | WASTE        | VY012203.D     | 16 Jan 2023 16:45 | Vial A                  | KP/MD    | Ok       |
| 12  | VY0116SBS01  | VY0116SBS01  | VY012204.D     | 16 Jan 2023 17:08 |                         | KP/MD    | Ok,M     |
| 13  | VY0116SBSD01 | VY0116SBSD01 | VY012205.D     | 16 Jan 2023 17:31 |                         | KP/MD    | Ok,M     |
| 14  | O1159-01     | TP-E         | VY012206.D     | 16 Jan 2023 17:55 | Vial A                  | KP/MD    | Ok       |
| 15  | O1160-01     | B-04-SB02    | VY012207.D     | 16 Jan 2023 18:18 | Vial A                  | KP/MD    | Ok       |
| 16  | O1125-01RE   | B-04-SB01RE  | VY012208.D     | 16 Jan 2023 18:42 | Vial B, ISTD Fail       | KP/MD    | Confirms |
| 17  | O1157-01     | TR-4-011323  | VY012209.D     | 16 Jan 2023 19:05 | Vial A, Not purged well | KP/MD    | Not Ok   |
| 18  | O1133-08     | TOT-VOC      | VY012210.D     | 16 Jan 2023 19:29 | Vial A                  | KP/MD    | Ok       |
| 19  | VSTDCCC050   | VSTDCCC050EC | VY012211.D     | 16 Jan 2023 19:51 |                         | KP/MD    | Ok,M     |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QC Batch ID # VY011623**

| Review By                | Krupa Patel                                           | Review On         | 1/17/2023 2:36:16 PM                      |
|--------------------------|-------------------------------------------------------|-------------------|-------------------------------------------|
| Supervise By             | Maresh Dadoda                                         | Supervise On      | 1/17/2023 2:38:16 PM                      |
| SubDirectory             | VY011623                                              | HP Acquire Method | MSVOA_Y HP Processing Method 82y011623s.m |
| STD. NAME                | STD REF.#                                             |                   |                                           |
| Tune/Reschk              | VP117988                                              |                   |                                           |
| Initial Calibration Stds | VP117989,VP117990,VP117991,VP117992,VP117993,VP117994 |                   |                                           |
| CCC                      | VP117996                                              |                   |                                           |
| Internal Standard/PEM    | VP116738                                              |                   |                                           |
| ICV/I.BLK                | VP117995                                              |                   |                                           |
| Surrogate Standard       |                                                       |                   |                                           |
| MS/MSD Standard          |                                                       |                   |                                           |
| LCS Standard             |                                                       |                   |                                           |

M : Manual Integration

Instrument ID: MSVOA\_Y

### Daily Analysis Runlog For Sequence/QC Batch ID # VY012023

|                                         |               |                   |                       |                      |              |  |  |
|-----------------------------------------|---------------|-------------------|-----------------------|----------------------|--------------|--|--|
| Review By                               | Mahesh Dadoda | Review On         | 1/25/2023 12:42:14 AM |                      |              |  |  |
| Supervise By                            | Krupa Patel   | Supervise On      | 1/25/2023 4:39:05 AM  |                      |              |  |  |
| SubDirectory                            | VY012023      | HP Acquire Method | MSVOA_Y               | HP Processing Method | 82y011623s.m |  |  |
| STD. NAME                               |               | STD REF.#         |                       |                      |              |  |  |
| Tune/Reschk<br>Initial Calibration Stds |               | VP118245          |                       |                      |              |  |  |
| CCC                                     |               | VP118243,VP118244 |                       |                      |              |  |  |
| Internal Standard/PEM                   |               | VP116738          |                       |                      |              |  |  |
| ICV/I.BLK                               |               |                   |                       |                      |              |  |  |
| Surrogate Standard                      |               |                   |                       |                      |              |  |  |
| MS/MSD Standard                         |               |                   |                       |                      |              |  |  |
| LCS Standard                            |               |                   |                       |                      |              |  |  |

| Sr# | SampleId     | ClientID     | Data File Name | Date-Time         | Comment                | Operator | Status   |
|-----|--------------|--------------|----------------|-------------------|------------------------|----------|----------|
| 1   | BFB          | BFB          | VY012279.D     | 20 Jan 2023 09:50 |                        | KP/MD    | Ok       |
| 2   | VSTDCCC050   | VSTDCCC050   | VY012280.D     | 20 Jan 2023 10:34 |                        | KP/MD    | Ok,M     |
| 3   | VY0120SBL01  | VY0120SBL01  | VY012281.D     | 20 Jan 2023 11:32 |                        | KP/MD    | Ok       |
| 4   | VY0120SBS01  | VY0120SBS01  | VY012282.D     | 20 Jan 2023 11:55 |                        | KP/MD    | Ok,M     |
| 5   | VY0120SBSD01 | VY0120SBSD01 | VY012283.D     | 20 Jan 2023 12:18 |                        | KP/MD    | Ok,M     |
| 6   | O1190-02     | PE-1D        | VY012284.D     | 20 Jan 2023 12:42 | Vial B                 | KP/MD    | Ok       |
| 7   | O1190-05RE   | PE-4RE       | VY012285.D     | 20 Jan 2023 13:05 | Vial B, Surrogate Fail | KP/MD    | Confirms |
| 8   | O1190-06     | PE-5         | VY012286.D     | 20 Jan 2023 13:29 | Vial B                 | KP/MD    | Ok       |
| 9   | O1190-09     | PE-8         | VY012287.D     | 20 Jan 2023 13:52 | Vial B                 | KP/MD    | Ok       |
| 10  | O1190-10     | PE-9         | VY012288.D     | 20 Jan 2023 14:16 | Vial B                 | KP/MD    | Ok       |
| 11  | O1190-11     | PE-10        | VY012289.D     | 20 Jan 2023 14:39 | Vial B                 | KP/MD    | Ok       |
| 12  | O1220-01     | MH-8         | VY012290.D     | 20 Jan 2023 15:02 | Vial A                 | KP/MD    | Ok       |
| 13  | O1220-04     | TP-G         | VY012291.D     | 20 Jan 2023 15:26 | Vial A                 | KP/MD    | Ok       |
| 14  | O1232-01     | B-P17A       | VY012292.D     | 20 Jan 2023 15:49 | Vial A                 | KP/MD    | Ok       |
| 15  | O1232-02     | B-P17B       | VY012293.D     | 20 Jan 2023 16:13 | Vial A                 | KP/MD    | Ok       |
| 16  | O1232-04     | B-P18A       | VY012294.D     | 20 Jan 2023 16:36 | Vial A, Not Purge      | KP/MD    | Not Ok   |
| 17  | O1232-05     | B-P18B       | VY012295.D     | 20 Jan 2023 17:00 | Vial A                 | KP/MD    | Ok       |
| 18  | O1232-06     | B-P19A       | VY012296.D     | 20 Jan 2023 17:23 | Vial A, ISTD Fail      | KP/MD    | ReRun    |
| 19  | O1232-07     | B-P19B       | VY012297.D     | 20 Jan 2023 17:46 | Vial A, ISTD Fail      | KP/MD    | Ok       |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QC Batch ID # VY012023**

| Review By                               | Mahesh Dadoda     | Review On         | 1/25/2023 12:42:14 AM                     |
|-----------------------------------------|-------------------|-------------------|-------------------------------------------|
| Supervise By                            | Krupa Patel       | Supervise On      | 1/25/2023 4:39:05 AM                      |
| SubDirectory                            | VY012023          | HP Acquire Method | MSVOA_Y HP Processing Method 82y011623s.m |
| STD. NAME                               | STD REF.#         |                   |                                           |
| Tune/Reschk<br>Initial Calibration Stds | VP118245          |                   |                                           |
| CCC                                     | VP118243,VP118244 |                   |                                           |
| Internal Standard/PEM                   | VP116738          |                   |                                           |
| ICV/I.BLK                               |                   |                   |                                           |
| Surrogate Standard                      |                   |                   |                                           |
| MS/MSD Standard                         |                   |                   |                                           |
| LCS Standard                            |                   |                   |                                           |

|    |            |              |            |                   |                   |       |        |
|----|------------|--------------|------------|-------------------|-------------------|-------|--------|
| 20 | O1232-08   | B-P35A       | VY012298.D | 20 Jan 2023 18:10 | Vial A            | KP/MD | Ok     |
| 21 | O1232-09   | B-P35B       | VY012299.D | 20 Jan 2023 18:33 | Vial A, ISTD Fail | KP/MD | ReRun  |
| 22 | O1232-10   | DUP01        | VY012300.D | 20 Jan 2023 18:57 | Vial A            | KP/MD | Ok     |
| 23 | O1244-02   | VNJ-223      | VY012301.D | 20 Jan 2023 19:20 | Vial A, Not Purge | KP/MD | Not Ok |
| 24 | O1252-01   | MH-9         | VY012302.D | 20 Jan 2023 19:43 | Vial A            | KP/MD | Ok     |
| 25 | O1239-01   | BORING-3     | VY012303.D | 20 Jan 2023 20:07 | Vial A            | KP/MD | Ok     |
| 26 | O1237-02   | 3A           | VY012304.D | 20 Jan 2023 20:30 | Vial A, Not Purge | KP/MD | Not Ok |
| 27 | VSTDCCC050 | VSTDCCC050EC | VY012305.D | 20 Jan 2023 20:53 |                   | KP/MD | Ok,M   |

M : Manual Integration



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA\_Y

## Daily Analysis Runlog For Sequence/QC Batch ID # VY012323

|                                         |  |                   |                   |  |         |                      |  |              |
|-----------------------------------------|--|-------------------|-------------------|--|---------|----------------------|--|--------------|
| Review By                               |  | Review On         |                   |  |         |                      |  |              |
| Supervise By                            |  | Supervise On      |                   |  |         |                      |  |              |
| SubDirectory                            |  | VY012323          | HP Acquire Method |  | MSVOA_Y | HP Processing Method |  | 82y011623s.m |
| STD. NAME                               |  | STD REF.#         |                   |  |         |                      |  |              |
| Tune/Reschk<br>Initial Calibration Stds |  | VP118120          |                   |  |         |                      |  |              |
| CCC                                     |  | VP118124,VP118125 |                   |  |         |                      |  |              |
| Internal Standard/PEM                   |  | VP116738          |                   |  |         |                      |  |              |
| ICV/I.BLK                               |  |                   |                   |  |         |                      |  |              |
| Surrogate Standard                      |  |                   |                   |  |         |                      |  |              |
| MS/MSD Standard                         |  |                   |                   |  |         |                      |  |              |
| LCS Standard                            |  |                   |                   |  |         |                      |  |              |

| Sr# | SampleId     | ClientID     | Data File Name | Date-Time         | Comment                                       | Operator | Status   |
|-----|--------------|--------------|----------------|-------------------|-----------------------------------------------|----------|----------|
| 1   | BFB          | BFB          | VY012306.D     | 23 Jan 2023 10:03 |                                               | KP/MD    | Ok       |
| 2   | VSTDCCC050   | VSTDCCC050   | VY012307.D     | 23 Jan 2023 12:09 |                                               | KP/MD    | Ok,NR    |
| 3   | VY0123SBL01  | VY0123SBL01  | VY012308.D     | 23 Jan 2023 13:04 |                                               | KP/MD    | Ok       |
| 4   | VY0123SBS01  | VY0123SBS01  | VY012309.D     | 23 Jan 2023 13:27 |                                               | KP/MD    | Ok,M     |
| 5   | VY0123SBSD01 | VY0123SBSD01 | VY012310.D     | 23 Jan 2023 13:50 |                                               | KP/MD    | Ok,M     |
| 6   | O1232-04     | B-P18A       | VY012311.D     | 23 Jan 2023 14:14 | Vial B                                        | KP/MD    | Ok       |
| 7   | O1232-06     | B-P19A       | VY012312.D     | 23 Jan 2023 14:37 | Vial B                                        | KP/MD    | Ok       |
| 8   | O1232-07     | B-P19B       | VY012313.D     | 23 Jan 2023 15:01 | Vial B, Not purged well                       | KP/MD    | Not Ok   |
| 9   | O1232-09RE   | B-P35BRE     | VY012314.D     | 23 Jan 2023 15:24 | Vial B, ISTD Fail and surr fail               | KP/MD    | Confirms |
| 10  | O1233-07     | B-P37B       | VY012315.D     | 23 Jan 2023 15:48 | Vial B, Internal Standard and Surrogate Fail  | KP/MD    | Ok       |
| 11  | O1233-08RE   | B-P38ARE     | VY012316.D     | 23 Jan 2023 16:11 | Vial B, Surrogate fail;Internal standard fail | KP/MD    | Confirms |
| 12  | O1244-02     | VNJ-223      | VY012317.D     | 23 Jan 2023 16:34 | Vial A                                        | KP/MD    | Ok       |
| 13  | O1251-05     | TOT-VOC-1    | VY012318.D     | 23 Jan 2023 16:58 | Vial A, ISTD Fail                             | KP/MD    | ReRun    |
| 14  | O1251-06     | TOT-VOC-2    | VY012319.D     | 23 Jan 2023 17:21 | Vial A                                        | KP/MD    | Ok       |
| 15  | O1251-11     | TOT-VOC-3    | VY012320.D     | 23 Jan 2023 17:45 | Vial A, ISTD Fail                             | KP/MD    | ReRun    |
| 16  | O1251-12     | TOT-VOC-4    | VY012321.D     | 23 Jan 2023 18:08 | Vial A, ISTD Fail                             | KP/MD    | ReRun    |
| 17  | O1237-02     | 3A           | VY012322.D     | 23 Jan 2023 18:31 | Vial B, Not Purge                             | KP/MD    | Not Ok   |
| 18  | O1254-01     | HR-04-012023 | VY012323.D     | 23 Jan 2023 18:55 | Vial A, ISTD Fail                             | KP/MD    | ReRun    |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QC Batch ID # VY012323**

|                                         |  |                   |  |                   |  |         |  |                      |  |              |  |
|-----------------------------------------|--|-------------------|--|-------------------|--|---------|--|----------------------|--|--------------|--|
| Review By                               |  | Review On         |  |                   |  |         |  |                      |  |              |  |
| Supervise By                            |  | Supervise On      |  |                   |  |         |  |                      |  |              |  |
| SubDirectory                            |  | VY012323          |  | HP Acquire Method |  | MSVOA_Y |  | HP Processing Method |  | 82y011623s.m |  |
| STD. NAME                               |  | STD REF.#         |  |                   |  |         |  |                      |  |              |  |
| Tune/Reschk<br>Initial Calibration Stds |  | VP118120          |  |                   |  |         |  |                      |  |              |  |
| CCC                                     |  | VP118124,VP118125 |  |                   |  |         |  |                      |  |              |  |
| Internal Standard/PEM                   |  | VP116738          |  |                   |  |         |  |                      |  |              |  |
| ICV/I.BLK                               |  |                   |  |                   |  |         |  |                      |  |              |  |
| Surrogate Standard                      |  |                   |  |                   |  |         |  |                      |  |              |  |
| MS/MSD Standard                         |  |                   |  |                   |  |         |  |                      |  |              |  |
| LCS Standard                            |  |                   |  |                   |  |         |  |                      |  |              |  |

|    |            |                      |            |                   |                        |       |       |
|----|------------|----------------------|------------|-------------------|------------------------|-------|-------|
| 19 | O1253-01   | SU-04-012023         | VY012324.D | 23 Jan 2023 19:18 | Vial A, Surrogate Fail | KP/MD | ReRun |
| 20 | O1243-01   | TJRC-11823-B4-0-10   | VY012325.D | 23 Jan 2023 19:42 | Vial A                 | KP/MD | Ok    |
| 21 | O1243-03   | TJRC-11823-B4-10-20  | VY012326.D | 23 Jan 2023 20:05 | Vial A, Surrogate Fail | KP/MD | ReRun |
| 22 | O1243-05   | TJRC-011923-B4-20-30 | VY012327.D | 23 Jan 2023 20:29 | Vial A                 | KP/MD | Ok    |
| 23 | O1243-07   | TJRC-011923-B4-30-40 | VY012328.D | 23 Jan 2023 20:52 | Vial A, ISTD Fail      | KP/MD | ReRun |
| 24 | VSTDCCC050 | VSTDCCC050EC         | VY012329.D | 23 Jan 2023 21:15 |                        | KP/MD | Ok,M  |

M : Manual Integration