



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: O1232

Client: Louis Berger U.S., Inc., A WSP Company

Analytical Method: SW8260D

Datafile : VY012309.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VY0123SBS01	Dichlorodifluoromethane	20	21.2	ug/Kg	106			64	136	
	Chloromethane	20	19.7	ug/Kg	99			70	130	
	Vinyl chloride	20	20.0	ug/Kg	100			72	129	
	Bromomethane	20	21.0	ug/Kg	105			58	141	
	Chloroethane	20	20.8	ug/Kg	104			69	130	
	Trichlorofluoromethane	20	21.2	ug/Kg	106			69	134	
	1,1,2-Trichlorotrifluoroethane	20	20.8	ug/Kg	104			81	123	
	1,1-Dichloroethene	20	20.6	ug/Kg	103			79	121	
	Acetone	100	87.9	ug/Kg	88			60	131	
	Carbon disulfide	20	20.2	ug/Kg	101			45	154	
	Methyl tert-butyl Ether	20	19.4	ug/Kg	97			77	129	
	Methyl Acetate	20	18.3	ug/Kg	92			69	149	
	Methylene Chloride	20	15.2	ug/Kg	76			39	175	
	trans-1,2-Dichloroethene	20	20.1	ug/Kg	101			80	123	
	1,1-Dichloroethane	20	20.5	ug/Kg	103			82	123	
	Cyclohexane	20	20.2	ug/Kg	101			76	122	
	2-Butanone	100	92.3	ug/Kg	92			69	131	
	Carbon Tetrachloride	20	20.7	ug/Kg	104			76	129	
	cis-1,2-Dichloroethene	20	20.5	ug/Kg	103			82	123	
	Bromochloromethane	20	19.2	ug/Kg	96			62	134	
	Chloroform	20	20.4	ug/Kg	102			82	125	
	1,1,1-Trichloroethane	20	20.7	ug/Kg	104			80	126	
	Methylcyclohexane	20	20.8	ug/Kg	104			77	123	
	Benzene	20	20.5	ug/Kg	103			84	121	
	1,2-Dichloroethane	20	20.2	ug/Kg	101			81	126	
	Trichloroethene	20	20.1	ug/Kg	101			83	122	
	1,2-Dichloropropane	20	20.3	ug/Kg	102			83	122	
	Bromodichloromethane	20	20.3	ug/Kg	102			82	123	
	4-Methyl-2-Pentanone	100	93.5	ug/Kg	94			70	135	
	Toluene	20	20.4	ug/Kg	102			83	122	
	t-1,3-Dichloropropene	20	19.6	ug/Kg	98			78	124	
	cis-1,3-Dichloropropene	20	20.1	ug/Kg	101			81	122	
	1,1,2-Trichloroethane	20	19.6	ug/Kg	98			82	125	
	2-Hexanone	100	92.0	ug/Kg	92			66	138	
	Dibromochloromethane	20	19.7	ug/Kg	99			79	125	
	1,2-Dibromoethane	20	19.3	ug/Kg	97			80	125	
	Tetrachloroethene	20	21.0	ug/Kg	105			83	125	
	Chlorobenzene	20	20.2	ug/Kg	101			84	122	
	Ethyl Benzene	20	20.3	ug/Kg	102			82	124	
	m/p-Xylenes	40	41.2	ug/Kg	103			83	124	
	o-Xylene	20	20.0	ug/Kg	100			83	123	
	Styrene	20	20.4	ug/Kg	102			82	124	
	Bromoform	20	19.1	ug/Kg	96			75	127	
	Isopropylbenzene	20	20.6	ug/Kg	103			82	124	
	1,1,2,2-Tetrachloroethane	20	19.3	ug/Kg	97			77	127	
	1,3-Dichlorobenzene	20	19.8	ug/Kg	99			83	122	
	1,4-Dichlorobenzene	20	19.5	ug/Kg	98			84	121	
	1,2-Dichlorobenzene	20	19.4	ug/Kg	97			83	124	
	1,2-Dibromo-3-Chloropropane	20	18.7	ug/Kg	94			66	134	



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: O1232

Client: Louis Berger U.S., Inc., A WSP Company

Analytical Method: SW8260D

Datafile : VY012309.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VY0123SBS01	1,2,4-Trichlorobenzene	20	17.9	ug/Kg	90			78	127	
	1,2,3-Trichlorobenzene	20	17.2	ug/Kg	86			70	137	