



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: louie01
Lab Code: CHEM Case No.: O1232 SAS No.: O1232 SDG NO.: O1232
Lab File ID: VY012280.D Date Analyzed: 01/20/2023
Instrument ID: MSVOA_Y Time Analyzed: 10:34
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	156820	13.428				
UPPER LIMIT	313640	13.928				
LOWER LIMIT	78410	12.928				
EPA SAMPLE NO.						
B-P17A	144924	13.43				
B-P17B	86164	13.43				
B-P18B	154191	13.43				
B-P19B	33154 *	13.43				
B-P35A	140878	13.43				
B-P35B	33340 *	13.43				
DUP01	114034	13.43				
VY0120SBL01	161157	13.43				
VY0120SBS01	155958	13.43				
VY0120SBSD01	155933	13.43				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.