



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH QC BATCHID: LB123785

	NA	NO	YES
1.Calibration Summary met criteria.	<u>          </u>	<u>          </u>	<u>  ✓  </u>
2.ICP Interference Check Sample Results Summary Submitted.	<u>          </u>	<u>          </u>	<u>  ✓  </u>
3. Serial Dilution Summary (if applicable) Submitted.	<u>          </u>	<u>          </u>	<u>  ✓  </u>
4.Laboratory Control Sample Summary (if applicable) Submitted.:	<u>          </u>	<u>          </u>	<u>  ✓  </u>
5.Blank Contamination - If yes, list compounds and concentrations in each blank:	<u>          </u>	<u>  ✓  </u>	<u>          </u>
6.Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	<u>          </u>	<u>          </u>	<u>  ✓  </u>
7.Sample Duplicate Analysis Met QC Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	<u>          </u>	<u>          </u>	<u>  ✓  </u>
8.Digestion Holding Time Met If not met, list number of days exceeded for each sample:	<u>          </u>	<u>          </u>	<u>  ✓  </u>
9.Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.	<u>          </u>	<u>          </u>	<u>  ✓  </u>

ADDITIONAL COMMENTS:

MS/MSD fail for O1251-01 (Sb,Cr,Cu,V) due to chemical interference during digestion process. Post spike recovery passed except Cr.

Analyst

Date

QA REVIEW

Date