CHEMTECH 284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Louis Berger U.S., Inc., A WSP Company **SDG No.:** O1233

Contract: loui01 **Lab Code:** CHEM **Case No.:** O1233 **SAS No.:** O1233

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.18	0.2	89	40 - 160	CV	01/24/2023	10:59	LB123775
CRI01	Arsenic	22.1	20.0	110	40 - 160	P	01/25/2023	12:24	LB123806
	Barium	109	100	109	40 - 160	P	01/25/2023	12:24	LB123806
	Cadmium	6.67	6.0	111	40 - 160	P	01/25/2023	12:24	LB123806
	Chromium	11.4	10.0	114	40 - 160	P	01/25/2023	12:24	LB123806
	Lead	12.3	12.0	102	40 - 160	P	01/25/2023	12:24	LB123806
	Selenium	20.5	20.0	103	40 - 160	P	01/25/2023	12:24	LB123806
	Silver	11.0	10.0	110	40 - 160	P	01/25/2023	12:24	LB123806