



Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Louis Berger U.S., Inc., A WSP Company **Level:** LOW **SDG No.:** O1233
Contract: lou01 **Lab Code:** CHEM **Case No.:** O1233 **SAS No.:** O1233
Matrix: Water **Sample ID:** O1255-02 **Client ID:** BORING-1DUP
Percent Solids for Sample: NA **Duplicate ID** O1255-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	447	J	435	J	3		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	16.7	J	15.1	J	10		P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P



Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Louis Berger U.S., Inc., A WSP Company **Level:** LOW **SDG No.:** O1233
Contract: lou01 **Lab Code:** CHEM **Case No.:** O1233 **SAS No.:** O1233
Matrix: Water **Sample ID:** O1255-02MS **Client ID:** BORING-1MSD
Percent Solids for Sample: NA **Duplicate ID** O1255-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3850		3860		0		P
Barium	ug/L	20	1360		1310		4		P
Cadmium	ug/L	20	970		965		1		P
Chromium	ug/L	20	1990		1950		2		P
Lead	ug/L	20	4750		4720		1		P
Mercury	ug/L	20	42.9		42.9		0		CV
Selenium	ug/L	20	9460		9440		0		P
Silver	ug/L	20	353		349		1		P