



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Louis Berger U.S., Inc., A WSP Company

SDG No.: O1233

Contract: loui01

Lab Code: CHEM

Case No.: O1233

SAS No.: O1233

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB150390BS Mercury	ug/L	4.0	3.88		97	80 - 120	CV



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Louis Berger U.S., Inc., A WSP Company

SDG No.: O1233

Contract: lou01

Lab Code: CHEM

Case No.: O1233

SAS No.: O1233

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB150432BS							
Arsenic	ug/L	4000	3980		100	80 - 120	P
Barium	ug/L	1000	981		98	80 - 120	P
Cadmium	ug/L	1000	971		97	80 - 120	P
Chromium	ug/L	2000	2050		102	80 - 120	P
Lead	ug/L	5000	4870		97	80 - 120	P
Selenium	ug/L	10000	10100		101	80 - 120	P
Silver	ug/L	380	379		100	80 - 120	P