



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: loui01
ProjectID: NYCDDC Phase II SCI Arthur Kill Road CEQR
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG No.: O1233

Calibration Sequence : FF123022		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	150435250	150435	FF011840.D
500	78485683	156971	FF011841.D
200	33452553	167263	FF011842.D
100	18725781	187258	FF011843.D
50	9861875	197238	FF011844.D
AVG RF : 171833		% RSD : 11.577	AVG RT : 15.0486