



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233

Continuing Calib Date: 01/20/2023 Initial Calibration Date(s): 01/17/2023 01/18/2023

Continuing Calib Time: 11:34 Initial Calibration Time(s): 19:14 02:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.51	5.51	5.41	5.61	0.00
Aroclor-1242-2 (2)	5.53	5.53	5.43	5.63	0.00
Aroclor-1242-3 (3)	5.59	5.59	5.49	5.69	0.00
Aroclor-1242-4 (4)	5.69	5.69	5.59	5.79	0.00
Aroclor-1242-5 (5)	6.43	6.43	6.33	6.53	0.00
Tetrachloro-m-xylene	4.33	4.33	4.23	4.43	0.00
Decachlorobiphenyl	10.10	10.10	10.00	10.20	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233

Continuing Calib Date: 01/20/2023 Initial Calibration Date(s): 01/17/2023 01/18/2023

Continuing Calib Time: 11:34 Initial Calibration Time(s): 19:14 02:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.67	4.67	4.57	4.77	0.00
Aroclor-1242-2 (2)	4.68	4.69	4.59	4.79	0.01
Aroclor-1242-3 (3)	4.86	4.86	4.76	4.96	0.00
Aroclor-1242-4 (4)	4.95	4.95	4.85	5.05	0.00
Aroclor-1242-5 (5)	5.47	5.47	5.37	5.57	0.00
Tetrachloro-m-xylene	3.58	3.58	3.48	3.68	0.00
Decachlorobiphenyl	8.60	8.60	8.50	8.70	0.00