



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233

Continuing Calib Date: 01/20/2023 Initial Calibration Date(s): 01/17/2023 01/18/2023

Continuing Calib Time: 18:26 Initial Calibration Time(s): 19:14 02:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.51	5.51	5.41	5.61	0.00
Aroclor-1016-2	(2)	5.53	5.53	5.43	5.63	0.00
Aroclor-1016-3	(3)	5.59	5.59	5.49	5.69	0.00
Aroclor-1016-4	(4)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-5	(5)	5.98	5.99	5.89	6.09	0.01
Aroclor-1260-1	(1)	7.12	7.12	7.02	7.22	0.00
Aroclor-1260-2	(2)	7.38	7.38	7.28	7.48	0.00
Aroclor-1260-3	(3)	7.74	7.74	7.64	7.84	0.00
Aroclor-1260-4	(4)	7.96	7.96	7.86	8.06	0.00
Aroclor-1260-5	(5)	8.28	8.28	8.18	8.38	0.00
Tetrachloro-m-xylene		4.33	4.33	4.23	4.43	0.00
Decachlorobiphenyl		10.10	10.09	9.99	10.19	-0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233

Continuing Calib Date: 01/20/2023 Initial Calibration Date(s): 01/17/2023 01/18/2023

Continuing Calib Time: 18:26 Initial Calibration Time(s): 19:14 02:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.67	4.67	4.57	4.77	0.00
Aroclor-1016-2	(2)	4.69	4.68	4.58	4.78	0.00
Aroclor-1016-3	(3)	4.86	4.86	4.76	4.96	0.00
Aroclor-1016-4	(4)	4.90	4.90	4.80	5.00	0.00
Aroclor-1016-5	(5)	5.12	5.12	5.02	5.22	0.00
Aroclor-1260-1	(1)	6.15	6.16	6.06	6.26	0.01
Aroclor-1260-2	(2)	6.34	6.34	6.24	6.44	0.00
Aroclor-1260-3	(3)	6.50	6.50	6.40	6.60	0.00
Aroclor-1260-4	(4)	6.97	6.97	6.87	7.07	0.00
Aroclor-1260-5	(5)	7.21	7.22	7.12	7.32	0.01
Tetrachloro-m-xylene		3.58	3.58	3.48	3.68	0.00
Decachlorobiphenyl		8.60	8.60	8.50	8.70	0.00