



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/17/2023 01/17/2023

Client Sample No.: CCAL01 Date Analyzed: 01/20/2023

Lab Sample No.: AR1660CCC500 Data File : PP054770.D Time Analyzed: 18:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.505	5.405	5.605	458.730	500.000	-8.3
Aroclor-1016-2	5.527	5.427	5.627	444.190	500.000	-11.2
Aroclor-1016-3	5.590	5.490	5.690	425.180	500.000	-15.0
Aroclor-1016-4	5.688	5.588	5.788	424.530	500.000	-15.1
Aroclor-1016-5	5.984	5.885	6.085	406.460	500.000	-18.7
Aroclor-1260-1	7.117	7.017	7.217	516.460	500.000	3.3
Aroclor-1260-2	7.375	7.275	7.475	523.750	500.000	4.8
Aroclor-1260-3	7.737	7.637	7.837	518.920	500.000	3.8
Aroclor-1260-4	7.962	7.863	8.063	512.200	500.000	2.4
Aroclor-1260-5	8.280	8.181	8.381	509.420	500.000	1.9
Decachlorobiphenyl	10.095	9.994	10.194	49.100	50.000	-1.8
Tetrachloro-m-xylene	4.332	4.231	4.431	53.290	50.000	6.6



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/17/2023 01/17/2023

Client Sample No.: CCAL01 Date Analyzed: 01/20/2023

Lab Sample No.: AR1660CCC500 Data File : PP054770.D Time Analyzed: 18:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.666	4.565	4.765	516.300	500.000	3.3
Aroclor-1016-2	4.685	4.584	4.784	518.240	500.000	3.6
Aroclor-1016-3	4.861	4.761	4.961	513.600	500.000	2.7
Aroclor-1016-4	4.903	4.804	5.004	471.600	500.000	-5.7
Aroclor-1016-5	5.116	5.017	5.217	540.130	500.000	8.0
Aroclor-1260-1	6.154	6.055	6.255	492.450	500.000	-1.5
Aroclor-1260-2	6.342	6.243	6.443	501.120	500.000	0.2
Aroclor-1260-3	6.497	6.398	6.598	480.660	500.000	-3.9
Aroclor-1260-4	6.971	6.872	7.072	476.120	500.000	-4.8
Aroclor-1260-5	7.214	7.115	7.315	484.720	500.000	-3.1
Decachlorobiphenyl	8.599	8.502	8.702	46.500	50.000	-7.0
Tetrachloro-m-xylene	3.579	3.478	3.678	49.930	50.000	-0.1