



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/17/2023 01/17/2023

Client Sample No.: CCAL03 Date Analyzed: 01/21/2023

Lab Sample No.: AR1660CCC500 Data File : PP054798.D Time Analyzed: 03:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.504	5.405	5.605	496.200	500.000	-0.8
Aroclor-1016-2	5.526	5.427	5.627	484.410	500.000	-3.1
Aroclor-1016-3	5.589	5.490	5.690	472.360	500.000	-5.5
Aroclor-1016-4	5.687	5.588	5.788	451.690	500.000	-9.7
Aroclor-1016-5	5.983	5.885	6.085	401.060	500.000	-19.8
Aroclor-1260-1	7.116	7.017	7.217	512.920	500.000	2.6
Aroclor-1260-2	7.374	7.275	7.475	515.760	500.000	3.2
Aroclor-1260-3	7.735	7.637	7.837	517.240	500.000	3.4
Aroclor-1260-4	7.962	7.863	8.063	514.740	500.000	2.9
Aroclor-1260-5	8.278	8.181	8.381	503.350	500.000	0.7
Decachlorobiphenyl	10.092	9.994	10.194	47.690	50.000	-4.6
Tetrachloro-m-xylene	4.332	4.231	4.431	51.030	50.000	2.1



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: loui01

Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/17/2023 01/17/2023

Client Sample No.: CCAL03 Date Analyzed: 01/21/2023

Lab Sample No.: AR1660CCC500 Data File : PP054798.D Time Analyzed: 03:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.664	4.565	4.765	493.300	500.000	-1.3
Aroclor-1016-2	4.683	4.584	4.784	488.460	500.000	-2.3
Aroclor-1016-3	4.860	4.761	4.961	483.740	500.000	-3.3
Aroclor-1016-4	4.902	4.804	5.004	461.100	500.000	-7.8
Aroclor-1016-5	5.115	5.017	5.217	489.830	500.000	-2.0
Aroclor-1260-1	6.152	6.055	6.255	488.060	500.000	-2.4
Aroclor-1260-2	6.341	6.243	6.443	511.050	500.000	2.2
Aroclor-1260-3	6.495	6.398	6.598	495.650	500.000	-0.9
Aroclor-1260-4	6.968	6.872	7.072	498.600	500.000	-0.3
Aroclor-1260-5	7.212	7.115	7.315	507.730	500.000	1.5
Decachlorobiphenyl	8.597	8.502	8.702	48.070	50.000	-3.9
Tetrachloro-m-xylene	3.579	3.478	3.678	47.060	50.000	-5.9