

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Report of Analysis

Client: Date Collected: Louis Berger U.S., Inc., A WSP Company Project: Date Received:

NYCDDC Phase II SCI Arthur Kill Road CEQR

Client Sample ID: B-P37B

Lab Sample ID: O1233-07

Analytical Method: SW8082A

Sample Wt/Vol: 30.04 Units: g

Soil Aliquot Vol: uL

Extraction Type:

1.0 GPC Factor: PH:

SW3541B Prep Method:

File ID/Qc Batch: Dilution: Prep Date Date Analyzed Prep Batch ID PP054791.D 1 01/20/23 10:35 01/21/23 00:28 PB150372

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.60	U	3.60	20.3	ug/kg
11104-28-2	Aroclor-1221	5.60	U	5.60	20.3	ug/kg
11141-16-5	Aroclor-1232	4.70	U	4.70	20.3	ug/kg
53469-21-9	Aroclor-1242	2.90	U	2.90	20.3	ug/kg
12672-29-6	Aroclor-1248	3.60	U	3.60	20.3	ug/kg
11097-69-1	Aroclor-1254	5.00	U	5.00	20.3	ug/kg
37324-23-5	Aroclor-1262	4.00	U	4.00	20.3	ug/kg
11100-14-4	Aroclor-1268	6.90	U	6.90	20.3	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.9		40 - 162	115%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.7		32 - 176	98%	SPK: 20

Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

01/18/23

01/19/23

O1233

SOIL

83.5

10000

PCB

Decanted:

иL

SDG No.:

Final Vol:

Injection Volume:

Test:

Matrix: % Solid:

() = Laboratory InHouse Limit