

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012023\
 Data File : PL080441.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jan 2023 23:28
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/23/2023
 Supervised By : Ankita Jodhani 01/23/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 06:18:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012023.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 20 17:01:11 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.352	2.690	108.0E6	111.3E6	50.346	50.674
28)	SA Decachlor...	8.790	7.838	84669230	112.1E6	48.362	47.061
Target Compounds							
2)	A alpha-BHC	3.800	3.191	164.6E6	163.9E6	51.967	51.535
3)	MA gamma-BHC...	4.129	3.519	153.1E6	152.9E6	52.065	51.175
4)	MA Heptachlor	4.713	3.860	151.7E6	160.5E6	51.318	50.433
5)	MB Aldrin	5.051	4.140	138.4E6	150.7E6	50.589	51.084
6)	B beta-BHC	4.325	3.815	67616599	66340945	49.535	49.707
7)	B delta-BHC	4.568	4.045	149.9E6	148.4E6	51.781	51.453
8)	B Heptachlo...	5.476	4.643	130.0E6	138.7E6	50.249	49.836
9)	A Endosulfan I	5.857	5.012	117.7E6	127.8E6	49.840	49.143
10)	B gamma-Chl...	5.729	4.894	127.3E6	140.0E6	50.419m	49.396
11)	B alpha-Chl...	5.810	4.958	126.4E6	140.7E6	50.065	49.132
12)	B 4,4'-DDE	5.987	5.150	112.2E6	118.4E6	50.362	50.189
13)	MA Dieldrin	6.132	5.277	120.0E6	134.9E6	50.144	49.863
14)	MA Endrin	6.359	5.552	108.4E6	117.2E6	50.109	50.509
15)	B Endosulfa...	6.578	5.846	98222849	94520874	42.849	45.361
16)	A 4,4'-DDD	6.501	5.704	91126693	95623128	49.742	50.885
17)	MA 4,4'-DDT	6.814	5.956	100.7E6	106.0E6	48.464	50.430
18)	B Endrin al...	6.707	6.026	80795772	87040737	48.572	47.665
19)	B Endosulfa...	6.941	6.249	102.7E6	113.7E6	49.597	48.909
20)	A Methoxychlor	7.292	6.534	54120755	59284897	49.054	48.856
21)	B Endrin ke...	7.421	6.754	108.5E6	125.3E6	49.165	49.067
22)	Mirex	7.893	6.938	84614334	108.9E6	49.200	47.814

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012023\
Data File : PL080441.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jan 2023 23:28
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 01/23/2023
Supervised By :Ankita Jodhani 01/23/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 06:18:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012023.M
Quant Title : GC Extractables
QLast Update : Fri Jan 20 17:01:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

