

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013123\
 Data File : BP013677.D
 Acq On : 31 Jan 2023 17:04
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP013123

Quant Time: Feb 01 04:11:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 01 03:43:48 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2	1,4-Dioxane	40.000	39.976	0.1	93	0.00
3	Pyridine	40.000	39.749	0.6	89	0.00
4	n-Nitrosodimethylamine	40.000	40.075	-0.2	92	0.00
5 S	2-Fluorophenol	80.000	82.665	-3.3	94	0.00
6	Aniline	40.000	41.110	-2.8	94	0.00
7 S	Phenol-d6	80.000	81.298	-1.6	93	0.00
8	2-Chlorophenol	40.000	40.739	-1.8	93	0.00
9	Benzaldehyde	40.000	42.434	-6.1	96	0.00
10 C	Phenol	40.000	40.611	-1.5	92	0.00
11	bis(2-Chloroethyl)ether	40.000	39.845	0.4	92	0.00
12	1,3-Dichlorobenzene	40.000	40.511	-1.3	93	0.00
13 C	1,4-Dichlorobenzene	40.000	40.671	-1.7	93	0.00
14	1,2-Dichlorobenzene	40.000	41.009	-2.5	94	0.00
15	Benzyl Alcohol	40.000	41.907	-4.8	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.641	-1.6	93	0.00
17	2-Methylphenol	40.000	41.432	-3.6	94	0.00
18	Hexachloroethane	40.000	40.863	-2.2	95	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.119	-5.3	94	0.00
20	3+4-Methylphenols	40.000	41.657	-4.1	94	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	94	0.00
22	Acetophenone	40.000	40.938	-2.3	94	0.00
23 S	Nitrobenzene-d5	80.000	86.363	-8.0	96	0.00
24	Nitrobenzene	40.000	42.614	-6.5	96	0.00
25	Isophorone	40.000	40.564	-1.4	93	0.00
26 C	2-Nitrophenol	40.000	39.244	1.9	96	0.00
27	2,4-Dimethylphenol	40.000	40.713	-1.8	92	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.798	0.5	93	0.00
29 C	2,4-Dichlorophenol	40.000	41.196	-3.0	95	0.00
30	1,2,4-Trichlorobenzene	40.000	41.108	-2.8	96	0.00
31	Naphthalene	40.000	39.863	0.3	93	0.00
32	Benzoic acid	40.000	39.559	1.1	96	0.00
33	4-Chloroaniline	40.000	40.768	-1.9	93	0.00
34 C	Hexachlorobutadiene	40.000	40.111	-0.3	95	0.00
35	Caprolactam	40.000	40.217	-0.5	93	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.148	-0.4	92	0.00
37	2-Methylnaphthalene	40.000	39.987	0.0	93	0.00
38	1-Methylnaphthalene	40.000	39.717	0.7	93	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	95	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.720	-1.8	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.631	-4.1	99	0.00
42 S	2,4,6-Tribromophenol	80.000	84.468	-5.6	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.869	-2.2	94	0.00
44	2,4,5-Trichlorophenol	40.000	41.409	-3.5	95	0.00
45 S	2-Fluorobiphenyl	80.000	80.872	-1.1	94	0.00
46	1,1'-Biphenyl	40.000	40.461	-1.2	94	0.00
47	2-Chloronaphthalene	40.000	40.788	-2.0	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013123\
 Data File : BP013677.D
 Acq On : 31 Jan 2023 17:04
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP013123

Quant Time: Feb 01 04:11:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 01 03:43:48 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.177	2.1	100	0.00
49	Acenaphthylene	40.000	40.679	-1.7	95	0.00
50	Dimethylphthalate	40.000	40.549	-1.4	95	0.00
51	2,6-Dinitrotoluene	40.000	40.124	-0.3	98	0.00
52 C	Acenaphthene	40.000	40.629	-1.6	96	0.00
53	3-Nitroaniline	40.000	40.284	-0.7	97	0.00
54 P	2,4-Dinitrophenol	40.000	38.845	2.9	100	0.00
55	Dibenzofuran	40.000	40.481	-1.2	96	0.00
56 P	4-Nitrophenol	40.000	43.049	-7.6	99	0.00
57	2,4-Dinitrotoluene	40.000	39.598	1.0	98	0.00
58	Fluorene	40.000	40.424	-1.1	95	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.741	-4.4	97	0.00
60	Diethylphthalate	40.000	40.521	-1.3	95	0.00
61	4-Chlorophenyl-phenylether	40.000	40.607	-1.5	96	0.00
62	4-Nitroaniline	40.000	40.916	-2.3	99	0.00
63	Azobenzene	40.000	40.807	-2.0	95	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.788	3.0	99	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.855	0.4	96	0.00
67	4-Bromophenyl-phenylether	40.000	40.538	-1.3	98	0.00
68	Hexachlorobenzene	40.000	40.353	-0.9	99	0.00
69	Atrazine	40.000	43.303	-8.3	99	0.00
70 C	Pentachlorophenol	40.000	42.900	-7.2	100	0.00
71	Phenanthrene	40.000	40.277	-0.7	97	0.00
72	Anthracene	40.000	40.729	-1.8	97	0.00
73	Carbazole	40.000	40.564	-1.4	97	0.00
74	Di-n-butylphthalate	40.000	40.958	-2.4	97	0.00
75 C	Fluoranthene	40.000	40.989	-2.5	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzidine	40.000	41.348	-3.4	95	0.00
78	Pyrene	40.000	40.075	-0.2	98	0.00
79 S	Terphenyl-d14	80.000	82.130	-2.7	101	0.00
80	Butylbenzylphthalate	40.000	41.894	-4.7	102	0.00
81	Benzo(a)anthracene	40.000	40.474	-1.2	100	0.00
82	3,3'-Dichlorobenzidine	40.000	41.710	-4.3	99	0.00
83	Chrysene	40.000	40.301	-0.8	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.578	-1.4	99	0.00
85 c	Di-n-octyl phthalate	40.000	40.522	-1.3	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.414	-1.0	101	0.00
88	Benzo(b)fluoranthene	40.000	41.018	-2.5	103	0.00
89	Benzo(k)fluoranthene	40.000	40.644	-1.6	98	0.00
90 C	Benzo(a)pyrene	40.000	40.298	-0.7	100	0.00
91	Dibenzo(a,h)anthracene	40.000	40.629	-1.6	101	0.00
92	Benzo(g,h,i)perylene	40.000	40.387	-1.0	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013123\
Data File : BP013677.D
Acq On : 31 Jan 2023 17:04
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP013123

Quant Time: Feb 01 04:11:36 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 01 03:43:48 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0