



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary
SW-846

SDG No.: O1233

Client: Louis Berger U.S., Inc., A WSP Company

Analytical Method: SW8270E

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits	
		Result	Result							High	RPD
Lab Sample ID:	O1233-02MS	Client Sample ID:	B-P34AMS					DataFile:	BF132171.D		
Naphthalene	2000	0	1600	ug/Kg	80				72	110	
Acenaphthylene	2000	0	1600	ug/Kg	80				79	118	
Acenaphthene	2000	0	1800	ug/Kg	90				70	121	
Fluorene	2000	0	1600	ug/Kg	80				68	116	
Phenanthrene	2000	0	1500	ug/Kg	75				72	113	
Anthracene	2000	0	1500	ug/Kg	75				62	124	
Fluoranthene	2000	0	1700	ug/Kg	85				59	125	
Pyrene	2000	0	1600	ug/Kg	80				52	128	
Benzo(a)anthracene	2000	0	1700	ug/Kg	85				71	114	
Chrysene	2000	0	1600	ug/Kg	80				57	121	
Benzo(b)fluoranthene	2000	0	1800	ug/Kg	90				67	121	
Benzo(k)fluoranthene	2000	0	1700	ug/Kg	85				74	114	
Benzo(a)pyrene	2000	0	1600	ug/Kg	80				70	142	
Indeno(1,2,3-cd)pyrene	2000	0	1500	ug/Kg	75				61	125	
Dibenz(a,h)anthracene	2000	0	1700	ug/Kg	85				67	130	
Benzo(g,h,i)perylene	2000	0	1500	ug/Kg	75				53	140	