



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233
EPA Sample No.: SSTDCCC040 Date Analyzed: 01/24/2023
Lab File ID: BP013571.D Time Analyzed: 15:13
Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	306718	8.163	1319960	10.99	845480	14.80
UPPER LIMIT	613436	8.663	2639920	11.493	1690960	15.298
LOWER LIMIT	153359	7.663	659980	10.493	422740	14.298
EPA SAMPLE NO.						
01 PB150378BL	291029	8.16	1295640	10.99	837908	14.80

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.